

AMENDMENT NO. 1 JULY 1980
TO
IS : 6960-1973 SPECIFICATION FOR CATHETERS, METAL, FEMALE
Alteration

(*Page 2, Clause 5.1*) — Substitute the following for the existing clause:

"5.1 *Corrosion Resistance Test (for Stainless Steel Only)* — Test the sample in accordance with IS : 7531-1975 ' Method for boiling and autoclaving test for corrosion resistance of stainless steel surgical instruments '. The sample shall show no sign of corrosion after the test. "

(CPDC 25)