

AMENDMENT NO. 1 JUNE 2021

TO

**IS 11552 : 2020 LIQUID NITROGEN VESSELS OF
CAPACITY UP TO 75 LITRES — SPECIFICATION**

(Second Revision)

(Page 5, clause 11.4.2) — Substitute the following for the existing:

‘After vibration tests, the vessels shall be again subjected to performance test laid down in 4.2; for seven days for vessels having capacity more than 1 litre and 3 days for vessels having capacity less than 1 litre and a deviation of 25 percent from the initial loss rate is acceptable. If one or more of these three vessels are not found satisfactory in performance testing, a fresh batch of ten vessels shall be manufactured and tested as above and all of them shall meet the requirements. The duration of performance test after vibration test shall be maximum 30 days.’