

AMENDMENT NO. 1 MAY 2025
TO
IS 5143 : 2024 ADJUSTABLE AXILLARY CRUTCHES — SPECIFICATION
(Third Revision)

(Page 2, clause 5.2) — Substitute the following for the existing:

5.2 Side bars of the aluminium crutches shall be powder-coated or shall be anodized as per Grade AC 10 of IS 1868.

(Page 4, Fig. 2) — Substitute ‘FOUR HOLES, PITCH 25’ *for* ‘THREE HOLES, PITCH 35’.

(MHD 09)

Price Group 1
Publications, BIS, New Delhi