



AMENDMENT NO. 1 NOVEMBER 1985

TO

IS:5181-1969 SPECIFICATION FOR PRESS,
FLASK, DENTAL

(Page 1, clause 2) - Add the following new matter under this clause:

'2.1 Tolerances on various dimensions shall be permitted as given below:

- a) ± 0.1 mm on dimensions up to 10.0 mm;
- b) ± 0.2 mm on dimensions above 10.0 mm and up to 20.0 mm;
- c) ± 0.5 mm on dimensions above 20.0 mm and up to 50.0 mm;
- d) ± 1.0 mm on dimensions above 50.0 mm and up to 100.0 mm;
- e) ± 2.0 mm on dimensions above 100.0 mm and up to 200.0 mm; and
- f) ± 2.5 mm on dimensions above 200.0 mm.'

(CPDC 21)

Reprography Unit, ISI, New Delhi, India