

AMENDMENT NO. 1 NOVEMBER 1992
TO
IS 6571 : 1991 REHABILITATION EQUIPMENT—
WHEELCHAIRS, NON-FOLDING, ADULT SIZE,
INSTITUTIONAL MODEL — SPECIFICATION

(First Revision)

(Page 1, Table 1, tenth row) — Substitute the following for the existing entry:

'Clearance of frame from floor 90, Min'

(Page 5, clause 9.3) — Insert the following new clause after 9.3:

'9.4 Stability Test

The wheelchair must resist toppling. All wheels must remain in contact with the surface of a 9° slope when the wheelchair is loaded with a 100 kg test load and positioned on the 9° slope with the front of the chair, pointed up-slope and the locks on the drive wheels engaged.'

(MHD10)