

AMENDMENT NO. 2 FEBRUARY 1993
TO
IS 6571:1991 REHABILITATION EQUIPMENT —
WHEELCHAIRS, NON-FOLDING, ADULT SIZE,
INSTITUTIONAL MODEL — SPECIFICATION

(First Revision)

(Page 3, clause 6.6) — Substitute the following for the existing clause:

"6.6 The solid tyres used in castors shall be made of non-marking resilient rubber with a minimum hardness of 60 to 70 shore 'A' with smooth treads."

(Page 3, clause 6.7.1, second sentence) — Substitute the following for the existing sentence:

"The solid tyres used in castors shall be made of non-marking resilient rubber with a minimum hardness of 60 to 70 shore 'A' with smooth treads."

(Page 4, clause 7.7.1) — Substitute the following for the existing clause:

"7.7.1 The wheelchair shall have two castors which shall be able to swivel through 360° in both directions, firmly secured to the frame in the front with either of the following combinations of bearings for smooth and silent performance:

- a) Two ball/thrust bearings, or
- b) One ball/thrust bearing and a pair of swivelling bush bearings.'

(Page 4, clause 7.7.3) — Substitute the following for the existing clause:

"7.7.3 The tyre shall be not less than 150 mm in diameter. The tyre shall be snap-on type and it shall not roll on the wheel rim under normal loads.'

(Page 4, clause 7.8) — Substitute the following for the existing clause:

"7.8 Wheels

The wheels shall be fixed to the frame in such a manner that the fitting shall be rugged enough to withstand the shocks during normal use. The wheels shall not rotate in more than one plane when the chair loaded with 100 kg is propelled. The wheels shall be removable from the chair without disturbing the bearing assembly.'

(MHD 10)