

AMENDMENT No. 3 APRIL 1995
TO
IS 6571 : 1991 REHABILITATION EQUIPMENT—
WHEELCHAIRS, NON-FOLDING, ADULT SIZE,
INSTITUTIONAL MODEL — SPECIFICATION

(*First Revision*)

(*Page 1, Table 1, eleventh row*)—Substitute the following for the existing entry:

‘Wheel diameter 40-534
 (24" × 1½")’

[*Page 3, clause 6.6 (see also Amendment No. 2)*] — Substitute the - following for the existing clause:

“**6.6** The solid tyres used in castors shall be made of non-marking resilient rubber with a minimum hardness of 60 shore ‘A’ with smooth treads.”

(*Page 3, clause 6.7.1, first sentence*) - Substitute the following for the existing sentence:

‘The rear wheels shall be of 40-534 (24" × 1½") size and shall have rims, tyres, tubes, nipples, spokes and washers as per the standard bicycle trade.’

[*Page 4, clause 7.7.3 (see also Amendment No. 2)*] — Substitute the following for the existing clause:

‘**7.7.3** The tyre shall be not less than 150 mm in diameter. The tyre shall be snap-on type and it shall not roll on the wheel rim under a load of 100 kg.’

(MHD10)