

AMENDMENT NO. 1 NOVEMBER 1984
TO

IS: 6581-1972 SPECIFICATION FOR ENDOTRACHEAL
TUBES (RUBBER)

Addenda

(*Page 4, clause 3.4*) - Add a new clause as follows:

'3.5 The endotracheal tubes shall be red in colour.'

(*Page 10, clause 8*) - Add new clause as follows:

'9. TEST

9.1 The air inflated pilot balloon and endotracheal tube shall not show any sign of leakage when tested under a pressure of 600 mm of water column.'

(CPDC 13)