



AMENDMENT NO. 1 OCTOBER 1965

TO

IS:6593-1972 SPECIFICATION FOR ELECTRIC
SEROLOGICAL WATER-BATHS

(Page 4, clause 5.2) - Substitute the following
for the existing first sentence:

'These shall be of copper, brass, stainless steel
or any other material suitable for the purpose.'

(CPDC 14)

Reprography Unit, ISI, New Delhi, India