

AMENDMENT NO. 3 AUGUST. 1998
TO
IS 6593 : 1972 SPECIFICATION FOR ELECTRIC
SEROLOGICAL WATER-BATHS

(*Page 3, clause 2.1, fourth line*) - Substitute 'ambient to 80°C' for '30 to 80°C'.

(MHD 14)

Reprography Unit, BIS, New Delhi India