

AMENDMENT NO. 1 AUGUST 2024
TO
IS 758 : 2023 MEDICAL TEXTILES — ABSORBENT COTTON GAUZE —
SPECIFICATION
(Fifth Revision)

(Page 1, clause 4.2) — Insert the following new clause after 4.2:

‘4.3 Cleanliness—Microbial/Bioburden Test

The cotton gauze shall confirm the requirement of cleanliness—microbial/bioburden for sterile and non-sterile product. For sterile product, ‘no viable microorganism shall be present’ and for non-sterile product, it shall be ≤ 300 (CFU/100 cm²) when tested in accordance with IS/ISO 11737 (Part 2) and IS/ISO 11737 (Part 1) respectively.’

[Page 2, Table 1, Sl No. (ix), col (4)] — Substitute ‘Viewing under ultraviolet light at 365 nm’ for ‘Viewing under ultraviolet light.’

[Page 2, Table 1, Sl No. (x)] — Delete.

(Page 4, Annex A) — Substitute ‘IS/ISO 11737-1 : 2018 Sterilization of health care products — Microbiological methods — Part 1: Determination of a population of microorganisms on products’ for ‘IS 10150 : 1981 Guide for sterilization of medical products’.

(Page 4, Annex A) — Insert the following before IS 14944 : 2020:

‘IS/ISO 11737-2 : 2019	Sterilization of health care products — Microbiological methods: Part 2 Tests of sterility performed in the definition, validation and maintenance of a sterilization process’.
------------------------	---