

AMENDMENT NO. 1 JULY 1980
TO
IS:8769-1978 METHODS OF DETERMINATION
OF ASH AND SULPHATED ASH

Corrigendum

*(Page 4, clause 5.2, line 7) - Substitute
'775 ± 25°C' for '775 ± 35°C'.*

(CDC 1)

Reprography Unit, ISI, New Delhi