

AMENDMENT NO. 1 JANUARY 2006
TO
IS 8897 : 1978 TABLES FOR CALIBRATION
AND METHOD OF VERIFICATION OF VOLUMETRIC
GLASSWARE

(*Page 3, Foreword, clause 0.4*) — Insert the following at the end of the clause:

‘The coefficient of cubical expansion is the increase in volume per unit volume for 1°C rise in temperature. For ordinary work it is assumed that the coefficient of cubical expansion is about three times the coefficient of linear expansion.’

(CHD 10)