


AMENDMENT NO. 1 FEBRUARY 2025
TO
IS 5912 : 1997 SPECIFICATION FOR ANEMOMETER, CUP COUNTER
(First Revision)

(Page 1, clause 6.1.2) — Substitute the following for the existing:

‘6.1.2 The three cups shall be fixed firmly to the central spider in a perfect vertical plane in such a way that they are well balanced with each other without showing any bias to any particular position (*see* Fig. 2). The mounting of the cups shall be firm enough to withstand even a wind speed of 200 km/h or more.



All dimensions in millimetres.

FIG. 2 SPIDER’.