

**AMENDMENT NO. 1 FEBRUARY 1996**  
**TO**  
**IS 6083 :1971 SPECIFICATION FOR TABLE,**  
**OBSTETRIC, LABOUR**

( Page 1, clause 4 ) — Substitute the following for the existing clause:

**'4. Manufacture** — The obstetric labour table shall have a welded framework of continuous lengths of angle iron of size 40 mm × 40 mm × 3 mm. The four upright members shall be connected at the floor end by a pair of longitudinal and a pair of transverse angles. The upright members shall support a formed metal top with perineal cut-out. The arrangement for welding the sheet metal tops and the uprights may be employed but the construction so obtained shall be sturdy and sound. On the perineal cut-out side the table shall have arrangement for receiving two lithotomy rods which shall be designed so as to provide rigid support. The table shall be free from burrs and sharp edges. The floor ends of the uprights shall be fitted with aluminium shoes. A drainage pipe shall be provided. The top of the table shall have sloping to facilitate drainage into the drain hole.'