



AMENDMENT NO. 1 SEPTEMBER 1984

TO

IS:6106-1971 SPECIFICATION FOR TABLES, OPERATION,
HYDRAULIC, MINOR

Alterations

(Page 4, clause 3.5) - Substitute the following
for the existing clause:

'3.5 Castor wheels shall be made from anti-static
nylon, anti-static rubber or steel. Castors shall
satisfy the requirements of IS:4034-1979†.

Note - When antistatic rubber or nylon wheels
are not available, the construction of the table
shall otherwise ensure transmission of static
charge to the earth.'

Addenda

(Page 4, clause 4.2, line 3) - Add the word
'preferably' between the words '(without mattress)'
and 'between 95 ± 5 cm'.

(Page 4, foot-note with '†' mark) - Add the
words '(first revision)' after the words 'hospital
equipment'.

(CPDC 14)

Reprography Unit, ISI, New Delhi, India