

AMENDMENT NO. 1 DECEMBER 2016
TO
IS 6190 : 1971 SPECIFICATION FOR ANAESTHETIC
FACE MASKS

(Page 3, clause 5.1) — Substitute the following for the existing clause:

“**5.1**The face mask shall have a female conical fitting conforming to IS 3234 (Part 2) : 1995 ‘Conical fittings with a 6 percent (LUER) taper for syringes, needles and certain other medical equipment: Part 2 Lock fitting (*second revision*)’.”

(Page 3, footnote *) — Delete.