

AMENDMENT NO. 1 MARCH 2023

TO

**IS 17243 : 2019 MEDICAL TEXTILES — TEST METHODS FOR COMPRESSES
FOR WOUND MANAGEMENT AND SURGICAL PROCEDURES**

(Page 4, Annex E, **E-3.2**) — Substitute the following for the existing: -

‘**E-3.2** Shake the container on the laboratory shaker set at 300 rpm to 350 rpm for 10 min.’