

AMENDMENT NO. 1 FEBRUARY 2024

TO

**IS 17354 : 2020 MEDICAL TEXTILES — DENTAL BIB / NAPKINS —
SPECIFICATION**

(Page 1, Table 1, Sl No. (iv), a)] — Substitute ‘Liquid absorbency time’ *for* ‘Sinking time’.

(Page 1, Table 1, Sl no. (iv), b)] — Substitute ‘Liquid absorptive capacity’ *for* ‘water holding capacity’.

(Page 1, Table 1) — Add the following new note at the end:

‘Artificial saliva (*see* ISO 20701) shall be used as test liquid if agreed between the buyer and seller.’

(Page 2, clause 7.1) — Insert the following information at the end:

‘f) Declared shelf-life of the product.’