

AMENDMENT NO. 1 FEBRUARY 2024
TO
IS 17508 : 2020 DISPOSABLE ADULT INCONTINENCE DIAPER —
SPECIFICATION

(Page 4, clause 7.6) — Insert the following new clause at the end:

‘7.7 Phthalate Test

The amount of phthalate present in adult incontinence diaper shall be < 0.1 percent (individual or in combination) when tested as per the method given in IS 9873 (Part 6). The phthalate test shall be done at raw material stage once for existing raw material and whenever there is a change in the raw material for manufacturing the product. The manufacturer of final product shall also do the phthalate test once in a year.’

(Page 3, clause 7.4, subtitle) — Substitute the following for the existing:

‘7.4 Biocompatibility Evaluation — Cytotoxicity, Irritation and Skin Sensitization (Optional)’

(Page 3, clause 7.4, first sentence) — Substitute the following for the existing:

‘If required by the buyer, the manufacturer shall ensure that raw material used for manufacturing the final product are safe for user based on its known toxicological characteristics at intended use.’

(Page 3, clause 7.4, second paragraph, second line) — Substitute ‘non-cytotoxic’ for ‘None’.

(Page 5, clause 9.1) — Insert the following at the end:

‘g) The information whether the material of the product is biocompatible that is, meets the requirement of the standard for biocompatibility evaluation – cytotoxicity, irritation and skin sensitization (if applicable).’