

**AMENDMENT NO. 1 JANUARY 1993
TO
IS 10110: 1982 SPECIFICATION FOR TABLE SPOON
TERMINAL DEVICE FOR ARTIFICIAL LIMBS**

(Page 1, Fig. 1):

- a) (*Length of stem*) — Substitute '75 to 80' for '75'.
- b) (*Diameter of stem*) — Substitute '8' for '7.5'.
- c) (*Length of thread*) — Substitute '15 to 20' for '17'.

(Page 1, clause 2.2) — Substitute the following for the existing clause:

'2.2 The end configuration size and angle may differ from Fig. 1, if agreed to between the purchaser and the supplier. The variation of the angle shall be 45° to 90°. However, the stem dimension and screw dimensions to fit the adaptor shall not vary.'

(Page 2, clause 3.2) — Insert the following new clause after 3.2:

'3.3 The adoption/stem may be of mild steel.'