

AMENDMENT NO. 2 SEPTEMBER 1994
TO
IS 10110:1982 SPECIFICATION FOR TABLE SPOON
TERMINAL DEVICE FOR ARTIFICIAL LIMBS
(*Page 2, clause 3.2*) — Substitute the following for the existing clause:
'3.2 If cast, material used shall be of non-corrosive grade of stainless steel.'
(*Page 2, clause 4.1*) — Substitute the word 'three' for 'two'.

(MHD 10)

Reprography Unit, BIS, New Delhi, India