

AMENDMENT NO. 1 JANUARY 1993
TO
IS 10111 : 1982 SPECIFICATION FOR TABLE FORK
TERMINAL DEVICE FOR ARTIFICIAL LIMBS

(Page 1 Fig.1)

a) (Length *of stem*) *Substitute* '75 to 80' for '75'

b) (*Diameter of stem*) — Substitute '8'for '7.5'.

c) (*Length of thread*) — Substitute '15 to 20'for '17'

(*Page* 1, clause 2.2) — Substitute the following for the existing clause

'22 The end configuration size and angle may differ from Fig 1, if agreed to between the purchaser and the supplier The variation of the angle shall be 45° to 90°. However, the stem dimension and screw dimensions to fit the adaptor shall not vary '

(*Page* 1, clause 3.2) — Insert the following new clause after 3 2

'3 3 The adoption/stem may be of mild steel '