

AMENDMENT NO. 2 SEPTEMBER 1994
TO
IS 10111: 1982 SPECIFICATION FOR TABLE FORK
TERMINAL DEVICE FOR ARTIFICIAL LIMBS

(Page 1, clause 3.2) = Substitute the following for the existing clause

'3.2 If cast, material used shall be of non-corrosive grade of stainless steel'

(Page 2, clause 4.1) — Substitute the word 'three' for 'two'