

AMENDMENT NO. 3 SEPTEMBER 1998
TO
IS 10150:1981 GUIDE FOR STERILIZATION OF
MEDICAL PRODUCTS

(*Page 4, clause 3.2.1.3, second line*) — Delete the word 'personnel'.

[*Page 5, clause 3.5(iv)*] — Substitute 'Posses' for 'Process'.

(*Page 6, clause 3.6, third sentence, seventh line*) — Substitute the following for the existing sentence:

'These discs or strips are placed alongside in the packets most difficult to sterilize in a sterilization chamber.'

(*Page 7, clause 4.1.4, second line*) — Insert 'ATCC No. 7953 or 12980/NCTC No. 10007' after '*Bacillus stearothermophilus*'.

(*Page 8, clause 4.2.1, second line*) — Insert 'ATCC No. 9372' after '*Bacillus subtilis*'.

(*Page 9, clause 4.4.1.1, third line*) — Substitute '1.33 MeV' for '1.33'.

(*Page 9, clause 4.4.1.1, fifth line*) — Substitute 'radioisotope' for 'isotope'.

(*Page 10, clause 4.4.4, last sentence*) — Substitute the following for the existing:

'Duration of radiation sterilization process depends upon the activity of the radioisotope source.'

(*Page 10, clause 4.4.4, Note, second line*) — Delete the word 'standard'.

(*Page 10, clause 4.4.5, first line*) — Substitute 'kGy' for 'KGy'.

(*Page 10, clause 4.4.6, second sentence*) — Substitute the following with the existing:

'Some suitable systems for routine dosimetry are:

- i) Dyed clear perspex,
- ii) Radiochromic dye films,
- iii) Ceric-cerous sulphate solution, and
- iv) Ethanol-cheorobenzene solution'.

Amend No. 3 to IS 10150:1981

(*Page 10, clause 4.4.8, second line*) — Insert '/ATCC No. 14884 or NCTC No. 8241' after 'ATCC No. 27142'.

(*Page 10, clause 4.4.9, ninth line*) — Insert 'surgical' after 'as'.

(*Page 17, clause B-2.1, sixth line*) — Insert '/ATCC No. 9372' after 'NCTC 10073'.

(MHD 14)