

AMENDMENT NO. 1 SEPTEMBER 2007
TO
IS 10245 (PART 1) : 1996 BREATHING APPARATUS
PART 1 CLOSED CIRCUIT BREATHING APPARATUS
(COMPRESSED OXYGEN CYLINDER) —
SPECIFICATION
(First Revision)

(Page 1, clause 3) — Insert the following new clause after 3.1.4 and renumber the subsequent clauses:

‘4 Classification

The Apparatus shall be classified as under:

Class of Apparatus		Nominal Working Duration, h	Minute Volume		
Without positive pressure	Positive pressure		Cycles/min	l/stroke	l/min
1 N	1 P	1	25	2.0	50.0
2 N	2 P	2	20	2.0	40.0
4 N	4 P	4	20	1.5	30.0

(Page 4, clause 4.8.2.1) — Substitute the following for the existing text:

‘The opening pressure of the moist relief valve at a constant flow of 1.0 l/min shall not exceed 10 mbar in the positive pressure sets and shall be between 1.0 mbar and 1.5 mbar in the sets without positive pressure.’

(Page 5, clause 4.8.2.2) — Substitute the following for the existing text:

‘The resistance of the relief valve shall not exceed 11 mbar in the positive pressure sets and shall not exceed 5.0 mbar in the sets without positive pressure when tested:

- a) at 50 l/min for sets with a continuous flow rate > 2 l/min.
- b) at 30 l/min for sets with a continuous flow rate up to 2 l/min.’

(Page 5, clause 4.10.1) — Substitute the following for the existing text:

‘The opening pressure of the lung governed supply mechanism measured at a constant flow of 10 l/min shall not be negative in the positive pressure apparatus and shall not exceed 3.5 mbar in the apparatus without positive pressure.’

(Page 7, clause 4.21) — Substitute the following for the existing text and table:

Minute Volume, l/min	Test time min	Measure of Breathing Resistance	
		Without positive pressure apparatus (mbar)	Positive pressure apparatus (mbar)
50 (25 cycles/min, 2.0 l/stroke)	60	-5 to 7	0 to 7
75 (30 cycles/min, 2.5 l/stroke)	5	-10 to 10	0 to 10

During the machine test at 100 l/min (40 cycles/min × 2.5 l/stroke) the apparatus shall function satisfactorily.

[Page 8, clause 6.2(b)] — Substitute the following for the existing:

‘Class as appropriate (see 4) and for temperature/humidity (see 4.20)’