

AMENDMENT NO.1 JULY 2007
IO
IS 9213 : 1979 SPECIFICATION FOR
BOD BOTTLES

(Page 4, clause 4.1.1) — Substitute the following for the existing clause

‘4.1.1 Limit of Alkalinity — When graded according to the method prescribed in IS 2303 (Part 1/Sec 1) 1994*, the glass of BOD bottles shall conform to Class HGB 1 ‘

(Page 4, footnote marked *) — Substitute the following for the existing

‘*Grading glass for alkalinity Part 1 Hydrolytic resistance, Section 1 Hydrolytic resistance of glass grains at 98°C Method of test and classification (first revision) ‘

(Page 6, clause 5.1.1) — Substitute the following for the existing clause

‘5.1.1 BIS Certification Mark

The bottles may also be marked with the Standard Mark

5.1.1.1 The use of the Standard Mark is governed by the provisions of *Bureau of Indian Standards Act, 1986* and the Rules and Regulations made thereunder. The details of conditions under which the licence for the use of Standard Mark may be granted to manufacturers or producers may be obtained from the Bureau of Indian Standards ‘

(CHD 10)