

**AMENDMENT NO. 1 APRIL 2012
TO
IS 9471 (PART 5) : 1980 SPECIFICATION FOR MODULAR
LOWER LIMB ORTHOTIC COMPONENTS**

PART 5 JOINT UNIT, KNEE

(Page 1, table under Fig. 1, Type II Drop-Lock) — Substitute the following dimensions for the existing:

Dimension	Size 1	Size 2	Size 3
<i>L</i> (± 0.5)	35	39	42
<i>L</i> 1 (± 0.5)	38	43	47
<i>W</i> 3 (± 0.25)	23	25	25

(Page 2, clause 3.4) — Add the following new clause after 3.4:

‘3.5 A rivet of minimum 4 mm diameter shall be used.’

(Page 2, clause 4.1) — Substitute the following for the existing clause:

“The male and female parts of the Joint unit shall be manufactured from high strength heat treatable stainless steel conforming to Designation 30Cr13 or 04Cr17Ni12Mo2 of IS 6911 : 1992 ‘Stainless steel plate, sheet and strip’.”

(Page 2, clause 4.2) — Substitute the following for the existing clause:

“The screws shall be manufactured from cold rolled bright steel annealed bars conforming to IS 9550 : 2001 ‘Bright steel bars — Specification’.”

(Page 2, clause 4.3) — Substitute the following for the existing clause:

“The nut shall be made from cold rolled bright steel annealed bars conforming to IS 9550 : 2001 ‘Bright steel bars — Specification’.”

(Page 2, clause 4.3) — Add the following new clause after 4.3:

‘4.4 The rivet shall be manufactured from a suitable aluminum alloy.’

(MHD 09)