

ZOOM ® View™ Report	Created 24 Apr 2017
---------------------	---------------------

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L311/269927 Last Office Interrogation

27 Mar 2017 Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (23 Mar 2017)

23 Apr 2017 19:17 RA Auto 23 Apr 2017 19:16 **RV** Auto

OK **Battery**

> Approximate time to explant: 9.5 years 100 min⁻¹ Magnet Rate

One Year Remaining Explant

		Explaint .		
Leads Data	Implant N/R N/R N/R	Previous Session	Most Recent	
Atrial				
Intrinsic Amplitude	N/R mV	2.2 mV	2.0 mV	
Pace Impedance	N/R Ω	615 Ω	619 Ω	
Pace Threshold	N/R V @ N/R ms	Auto 1.7 V @ 0.4 ms	Auto 0.9 V @ 0.4 ms	
Ventricular				
Intrinsic Amplitude	N/R mV	14.8 mV	16.9 mV @ 50 min ⁻¹	
Pace Impedance	N/R Ω	687 Ω	733 Ω	
Pace Threshold	N/R V @ N/R ms	Auto 0.7 V @ 0.4 ms	Auto 0.5 V @ 0.4 ms	
Prody Countary Since Last Box	ot (22 Mar 2017)			

Brady Counters Since Last Reset (23 Mar 2017)

Counters

Atrial 91% Paced 100% Paced Ventricular

Atrial Arrhythmia

% AT/AF 0

Settings

Ventricular Tachy Settings		Atrial Tachy Settings	
Ventricular Tachy EGM Storage	On	ATR Mode Switch 170 min ⁻¹	

AIR Mode Switch 170 min⁻¹ DDI Ventricular Tachy EGM Storage

160 min⁻¹ **Detection Rate**

Brady Settings

DDDR Pacing Output Mode

RYTHMIQ™ **AAIR With** Atrial Auto 2.0 V @ 0.4 ms Trend 2.5 V @ 0.4 ms VVI Backup Ventricular

Lower Rate Limit 55 min⁻¹ Sensitivity

Maximum Tracking Rate 130 min⁻¹ Fixed 0.5 mV Atrial Maximum Sensor Rate 130 min⁻¹ Ventricular Fixed 2.0 mV

Paced AV Delay 80 - 180 ms Leads Configuration (Pace/Sense)

Sensed AV Delay 65 - 150 ms Atrial **Bipolar** 240 - 280 ms Ventricular A-Refractory (PVARP) Bipolar

V-Refractory (VRP) 230 - 250 ms

2869 Software Version: 2.02 L311 Firmware Version: F_v1.00.00

© 2014 Boston Scientific Corporation or its affiliates. All rights reserved. Page 1 of 1

Clinician Signature: