

ZOOM ® View™ Report Created 04 Jun 2018

QUICK NOTES ® Report

Hucík, Stepán

Date of Birth

Device

7 Jan 1944

INCEPTA CRT-D P162/119874

Tachy Mode Monitor + Therapy

Last Office Interrogation

19 Jun 2017 Implant Date 12 Mar 2013

My Alerts

△ 04 Jun 2018 09:53 Check Atrial and LV Leads

Events Since Last Reset (14 Mar 2013)

30 Aug 2014 13:14 VT at 200 min⁻¹, **②** ATPx4, 17J 09 Aug 2014 16:46 VT at 181 min⁻¹, **②** ATPx4, 17J

See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 3.5 years Charge Time 10.7 s

Last Capacitor Re-form 27 Apr 2018 11:27

One Year Remaining

Leads Data Implant Previous Session Most Recent

12 Mar 2013

12 Mai 2013		
N/R mV	N/R mV	N/R mV
N/R Ω	>2500 Ω	>2500 Ω
N/R V @ N/R ms	4.0 V @ 0.4 ms	
7.0 mV	PVC mV	6.0 mV @ 44 min ⁻¹
$650~\Omega$	550 Ω	507 Ω
0.5 V @ 0.4 ms	0.7 V @ 0.4 ms	1.1 V @ 0.4 ms
12.0 mV	PVC mV	20.7 mV
$650~\Omega$	544 Ω	545 Ω
1.3 V @ 0.4 ms	1.1 V @ 0.4 ms	1.2 V @ 0.4 ms
45 Ω	56 Ω	52 Ω
	$\begin{array}{c} \text{N/R mV} \\ \text{N/R } \Omega \\ \text{N/R V @ N/R ms} \\ \\ 7.0 \text{ mV} \\ 650 \ \Omega \\ 0.5 \text{ V @ 0.4 ms} \\ \\ 12.0 \text{ mV} \\ 650 \ \Omega \\ \\ 1.3 \text{ V @ 0.4 ms} \\ \end{array}$	$\begin{array}{cccccccccccccccccccccccccccccccccccc$

Brady/CRT Counters Since Last Reset (14 Mar 2013)

Counters

Atrial 0% Paced Right Ventricular 90% Paced Left Ventricular 97% Paced

Atrial Arrhythmia

% AT/AF 0

(One Year of Data)

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 31J, 41J, 41Jx6 VT 205 min⁻¹ Scan Ramp 31J, 41J, 41Jx4 VT-1 160 min⁻¹ Scan Ramp 41J, 41J, 41Jx3

Brady/CRT Settings

Mode VVIR Pacing Output

Lower Rate Limit 60 min⁻¹ Right Ventricular

Maximum Sensor Rate 130 min⁻¹ Left Ventricular RV-Refractory (RVRP) 230 - 250 ms Sensitivity LV-Refractory (LVRP) 250 ms Atrial

Ventricular Pacing Chamber BiV Right Ventricular

2868 Software Version: 4.04

P162 Firmware Version: B_v1.02.00(4.01)

Boston Scient or its affiliates

© 2016

Boston Scientific Corporation or its affiliates. All rights reserved.

Page 1 of 2

Clinician Signature:

3.0 V @ 0.4 ms

3.0 V @ 0.4 ms

AGC 0.25 mV

AGC 0.6 mV

Settings (Continued)

Brady/CRT Settings

LV Offset 0 ms Sensitivity

Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Atrial Bipolar/Off
Right Ventricular Bipolar
Left Ventricular LVring>>RV / LVtip>>LVring

All Events Since Last Reset (14 Mar 2013) 25 Apr 2018 08:08 NonSustV at 169 min⁻¹, Nonsustained 20 Apr 2018 18:38 NonSustV at 153 min⁻¹, Nonsustained NonSustV at 156 min⁻¹, Nonsustained 14 Apr 2018 18:31 14 Apr 2018 17:08 NonSustV at 159 min⁻¹, Nonsustained NonSustV at 159 min⁻¹, Nonsustained 14 Apr 2018 17:02 14 Apr 2018 17:01 NonSustV at 158 min⁻¹, Nonsustained NonSustV at 161 min⁻¹, Nonsustained 14 Apr 2018 17:00 14 Apr 2018 16:55 NonSustV at 164 min⁻¹, Nonsustained 14 Apr 2018 16:54 NonSustV at 157 min⁻¹, Nonsustained 14 Apr 2018 14:00 NonSustV at 161 min⁻¹, Nonsustained 03 Dec 2015 12:23 VT-1 at 181 min⁻¹, No Therapy VT-1 at 172 min⁻¹, No Therapy 03 Dec 2015 12:11 03 Dec 2015 11:32 VT-1 at 146 min⁻¹, No Therapy 19 Oct 2015 14:51 VT-1 at 175 min⁻¹, No Therapy 23 Sep 2015 16:09 VT-1 at 163 min⁻¹, ATPx1 VT-1 at 173 min⁻¹, ATPx1 23 Sep 2015 15:54 23 Sep 2015 14:11 VT-1 at 157 min⁻¹, ATPx1 VT-1 at 162 min⁻¹, **♦** ATPx4 23 Sep 2015 13:11 VT-1 at 179 min⁻¹, ATPx5 23 Sep 2015 13:10 VT-1 at 173 min⁻¹. ◆ ATPx2 23 Sep 2015 13:07 VT-1 at 178 min⁻¹, ATPx1 21 Sep 2015 20:26 21 Sep 2015 19:19 VT-1 at 186 min⁻¹, ATPx1 VT-1 at 181 min⁻¹, ATPx8, 41J 17 Sep 2015 12:20 17 Sep 2015 12:18 VT-1 at 181 min⁻¹, ATPx5 17 Sep 2015 10:44 VT-1 at 196 min⁻¹, ATPx1 30 Aug 2015 17:02 VT-1 at 155 min⁻¹, **♦** ATPx3 VT-1 at 165 min⁻¹, No Therapy 30 Aug 2015 16:19 29 Aug 2015 14:33 VT-1 at 170 min⁻¹, No Therapy 29 Aug 2015 14:32 VT-1 at 186 min⁻¹. ATPx1 29 Aug 2015 14:20 VT-1 at 176 min⁻¹, ATPx1 18 May 2015 10:09 VT-1 at 180 min⁻¹, No Therapy VT-1 at 183 min⁻¹, **♦** ATPx8, 41J 25 Oct 2014 11:49 VT-1 at 155 min⁻¹, ATPx1 25 Oct 2014 11:38 VT-1 at 168 min⁻¹, **♦** ATPx8, 41J 10 Oct 2014 12:13 VT at 200 min⁻¹, **♦** ATPx4, 17J 30 Aug 2014 13:14 09 Aug 2014 16:46 VT at 181 min⁻¹, ATPx4, 17J VT at 178 min⁻¹, ATPx2 16 May 2013 17:54 VT at 179 min⁻¹, **♦** ATPx4, 17J 17 Apr 2013 15:14

2868 Software Version: 4.04 P162 Firmware Version: B_v1.02.00(4.01)

102 1 IIIIWale Velololi. B_V1.02.00(4.01)

© 2016

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: