



ZOOM® View™

QUICK NOTES® Report

Hlucík, Stepán

Date of Birth 7 Jan 1944
Device INCEPTA CRT-D P162/119874
Tachy Mode Monitor + Therapy

Report Created 06 Jun 2016

Last Office Interrogation
12 Oct 2015
Implant Date
12 Mar 2013

My Alerts

⚠ 06 Jun 2016 13:55 Check Atrial and LV Leads

Events Since Last Reset (14 Mar 2013)

30 Aug 2014 13:14 VT at 200 min⁻¹, ♦ ATPx4, 17J

09 Aug 2014 16:46 VT at 181 min⁻¹, ♦ ATPx4, 17J

See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 5 years

Charge Time 9.9 s

Last Capacitor Re-form 26 Apr 2016 09:20

One Year Remaining

Explant



Leads Data

Implant

Previous Session

Most Recent

12 Mar 2013

Atrial

Intrinsic Amplitude N/R mV N/R mV N/R mV

Pace Impedance N/R Ω >2500 Ω >2500 Ω

Pace Threshold N/R V @ N/R ms 4.0 V @ 0.4 ms

Right Ventricular

Intrinsic Amplitude 7.0 mV 8.1 mV 8.0 mV @ 71 min⁻¹

Pace Impedance 650 Ω 518 Ω 556 Ω

Pace Threshold 0.5 V @ 0.4 ms 1.1 V @ 0.4 ms 1.3 V @ 0.4 ms

Left Ventricular

Intrinsic Amplitude 12.0 mV 13.7 mV 19.0 mV

Pace Impedance 650 Ω 512 Ω 532 Ω

Pace Threshold 1.3 V @ 0.4 ms 1.3 V @ 0.4 ms 1.4 V @ 0.4 ms

Shock Vector

Shock Impedance 45 Ω 55 Ω 57 Ω

Brady/CRT Counters Since Last Reset (14 Mar 2013)

Atrial 0% Paced

Right Ventricular 91% Paced

Left Ventricular 98% Paced

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 31J, 41J, 41Jx6

VT 205 min⁻¹ Scan Ramp 31J, 41J, 41Jx4

VT-1 160 min⁻¹ Scan Ramp 41J, 41J, 41Jx3

Brady/CRT Settings

Mode VVIR

Lower Rate Limit 60 min⁻¹

Maximum Sensor Rate 130 min⁻¹

RV-Refractory (RVRP) 230 - 250 ms

LV-Refractory (LVRP) 250 ms

Ventricular Pacing Chamber BiV

LV Offset 0 ms

Pacing Output

Right Ventricular 3.0 V @ 0.4 ms

Left Ventricular 3.0 V @ 0.4 ms

Sensitivity

Atrial AGC 0.25 mV

Right Ventricular AGC 0.6 mV

Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Atrial Bipolar/Off
Right Ventricular Bipolar

Clinician Signature:

Settings (Continued)

Brady/CRT Settings

Leads Configuration (Pace/Sense)

Left Ventricular

LVring>>RV / LVtip>>LVring

All Events Since Last Reset (14 Mar 2013)

03 Jun 2016 16:58	NonSustV at 159 min ⁻¹ , Nonsustained
03 Jun 2016 15:21	NonSustV at 155 min ⁻¹ , Nonsustained
03 Jun 2016 15:19	NonSustV at 153 min ⁻¹ , Nonsustained
03 Jun 2016 15:15	NonSustV at 157 min ⁻¹ , Nonsustained
03 Jun 2016 15:15	NonSustV at 156 min ⁻¹ , Nonsustained
03 Jun 2016 15:14	NonSustV at 159 min ⁻¹ , Nonsustained
03 Jun 2016 15:07	NonSustV at 161 min ⁻¹ , Nonsustained
03 Jun 2016 15:02	NonSustV at 160 min ⁻¹ , Nonsustained
03 Jun 2016 15:00	NonSustV at 155 min ⁻¹ , Nonsustained
03 Jun 2016 14:59	NonSustV at 159 min ⁻¹ , Nonsustained
03 Dec 2015 12:23	VT-1 at 181 min ⁻¹ , No Therapy
03 Dec 2015 12:11	VT-1 at 172 min ⁻¹ , No Therapy
03 Dec 2015 11:32	VT-1 at 146 min ⁻¹ , No Therapy
19 Oct 2015 14:51	VT-1 at 175 min ⁻¹ , No Therapy
23 Sep 2015 16:09	VT-1 at 163 min ⁻¹ , ATPx1
23 Sep 2015 15:54	VT-1 at 173 min ⁻¹ , ATPx1
23 Sep 2015 14:11	VT-1 at 157 min ⁻¹ , ATPx1
23 Sep 2015 13:11	VT-1 at 162 min ⁻¹ , ♦ ATPx4
23 Sep 2015 13:10	VT-1 at 179 min ⁻¹ , ♦ ATPx5
23 Sep 2015 13:07	VT-1 at 173 min ⁻¹ , ♦ ATPx2
21 Sep 2015 20:26	VT-1 at 178 min ⁻¹ , ATPx1
21 Sep 2015 19:19	VT-1 at 186 min ⁻¹ , ATPx1
17 Sep 2015 12:20	VT-1 at 181 min ⁻¹ , ♦ ATPx8, 41J
17 Sep 2015 12:18	VT-1 at 181 min ⁻¹ , ♦ ATPx5
17 Sep 2015 10:44	VT-1 at 196 min ⁻¹ , ATPx1
30 Aug 2015 17:02	VT-1 at 155 min ⁻¹ , ♦ ATPx3
30 Aug 2015 16:19	VT-1 at 165 min ⁻¹ , No Therapy
29 Aug 2015 14:33	VT-1 at 170 min ⁻¹ , No Therapy
29 Aug 2015 14:32	VT-1 at 186 min ⁻¹ , ATPx1
29 Aug 2015 14:20	VT-1 at 176 min ⁻¹ , ATPx1
18 May 2015 10:09	VT-1 at 180 min ⁻¹ , No Therapy
25 Oct 2014 11:49	VT-1 at 183 min ⁻¹ , ♦ ATPx8, 41J
25 Oct 2014 11:38	VT-1 at 155 min ⁻¹ , ATPx1
10 Oct 2014 12:13	VT-1 at 168 min ⁻¹ , ♦ ATPx8, 41J
30 Aug 2014 13:14	VT at 200 min ⁻¹ , ♦ ATPx4, 17J
09 Aug 2014 16:46	VT at 181 min ⁻¹ , ♦ ATPx4, 17J
16 May 2013 17:54	VT at 178 min ⁻¹ , ♦ ATPx2
17 Apr 2013 15:14	VT at 179 min ⁻¹ , ♦ ATPx4, 17J