

Device Settings Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J276/102818

Last Office Interrogation

19 Sep 2016 Implant Date N/R

Prograr	

Last Programmed 19 Sep 2016

Ventricular Tachy

2869 Software Version: 2.02

J276 Firmware Version: D_v1.02.00(1.01)

Ventricular Tachy EGM Storage On

Detection Rate 160 min⁻¹ (375 ms)

Atrial Tachy

Therapy

ATR Mode Switch Details Atrial Tachy Response

ATR Mode Switch On Atrial Flutter Response On
Trigger Rate 170 min⁻¹ Trigger Rate 170 min⁻¹
Duration 8 cycles PMT Termination On

Entry Count 8 cycles Atrial Overdrive

Exit Count 8 cycles APP/ProACt Off

Fallback DDI

Time 00:30 mm:ss
ATR Fallback LRL 70 min⁻¹
Ventricular Rate Regulation On
Maximum Pacing Rate 130 min⁻¹

© 2014 Clinician Signature:

5 .						
Brady						
Settings		_				
Mode	DDDR	Output				
RYTHMIQ™	AAIR With VVI	●A	Auto 2.0 V @ 0.4 ms			
	Backup	■V	2.3 V @ 0.4 ms			
Lower Rate Limit	55 min ^{−1}	Sensitivity				
Maximum Tracking Rate	130 min ⁻¹	●A	Fixed 0.5 mV			
Maximum Sensor Rate	130 min ⁻¹	■V	Fixed 2.0 mV			
Paced AV Delay	80 - 180 ms	Leads				
Sensed AV Delay	65 - 150 ms	●A				
A-Refractory (PVARP)	240 - 280 ms	Pace	Bipolar			
V-Refractory (VRP)	230 - 250 ms	Sense	Bipolar			
PVARP after PVC	400 ms	Safety Switch	On			
AV Search +	On	■V				
Search AV Delay	300 ms	Pace	Bipolar			
Search Interval	32 cycles	Sense	Bipolar			
Blanking	·	Safety Switch	On			
A-Blank after V-Pace	125 ms	Rate Adaptive Pacing				
A-Blank after V-Sense	45 ms	Minute Ventilation	On			
V-Blank after A-Pace	65 ms	Response Factor	8			
Magnet Response	Pace Async	Fitness Level	Active			
Noise Response	DOO	Ventilatory Threshold	120 min ⁻¹			
Rate Enhancements		Ventilatory Thresh. Response	70 %			
Rate Smoothing		Accelerometer	Passive			
Up	Off %					
Down	Off %					
Sudden Brady Response	Off					
Setup						
Telemetry						
Enable use of ZIP™ telemetry		On				
Trending Setup						
Recording Method	30	30 Second Average				
Duration		25 hours				
		<u> </u>				

Data Storage Continuous

Sleep Schedule

Sleep Start Time 23:00 hh:mm Sleep Duration 07 hours

Leads Status Setup

		Daily Intrinsic Amplitude	Daily Impedance	Impedance Low F	Limits ligh	Pace Threshold
● A	Pace/Sense	On On	On	200 - 2		On O"
■V	Pace/Sense	On	On	200 - 2	000 02	Off

2869 Software Version: 2.02 J276 Firmware Version: D_v1.02.00(1.01)

© 2014 Boston Scientific Corporation or its affiliates. All rights reserved. Page 2 of 2

Clinician Signature: