

Combined Follow-up Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J275/101893

Last Office Interrogation

09 Jan 2017 Implant Date **N/R**

My Alerts

▲ 08 Jan 2018 10:41 Check Lead

Events Since Last Reset (02 Jun 2014)

14 Dec 2017 11:09 Tachy at 177 min⁻¹
14 Dec 2017 11:08 Tachy at 181 min⁻¹
See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 8.5 years Magnet Rate 100 min⁻¹

One Year Remaining

Explant Most Recen

Leads Data	Implant N/R N/R N/R	Previous Session	Most Recent
Intrinsic Amplitude	N/R mV	>25.0 mV	20.5 mV @ 39 min ⁻¹
Pace Impedance	N/R Ω	588 Ω	576 Ω
Pace Threshold	N/R V @ N/R ms	0.6 V @ 0.4 ms	Auto 1.0 V @ 0.4 ms

Settings

Tachy Settings

Tachy EGM Storage On
Detection Rate 160 min⁻¹

Brady Settings

Mode VVIR Pacing Output Lower Rate Limit 55 min⁻¹

Maximum Sensor Rate 115 min⁻¹ Sensitivity

Refractory 250 - 250 ms Fixed 2.0 mV

Leads Configuration (Pace/Sense)

Unipolar/Bipolar

Auto 1.5 V @ 0.4 ms

Rate Adaptive Pacing

Minute Ventilation On Accelerometer Passive

2869 Software Version: 2.02


J275 Firmware Version: D_v1.02.00(1.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 1 of 5

Clinician Signature:




Tachy Counters	Since Last Reset 02 Jun 2014	Device Totals	
Episode Counters			
Total Episodes	63.3K	63.3K	
Tachy Episodes	1.7K	1.7K	
Nonsustained Episodes	61.7K	61.7K	
Brady Counters	Reset Before Last N/R day(s) N/R to 02 Jun 2014	Since Last Reset 1316 day(s) 02 Jun 2014 to Today	
Counters		·	
% Paced	0	52	
Intrinsic Promotion			
Rate Hysteresis			
% Successful	0	0	

Histograms Reset Before Last (N/R)



Histograms Since Last Reset (02 Jun 2014)



All Events Since Last Reset (02 Jun 2014)

LVCIII3 OIIICC Last Nesct	(02 0411 20 1 7)
04 Jan 2018 16:20	NonSust at 134 min ⁻¹
29 Dec 2017 13:05	NonSust at 137 min ⁻¹
29 Dec 2017 12:59	NonSust at 151 min ⁻¹
29 Dec 2017 12:56	NonSust at 141 min ⁻¹
29 Dec 2017 12:55	NonSust at 141 min ⁻¹
29 Dec 2017 12:55	NonSust at 141 min ⁻¹
29 Dec 2017 12:54	NonSust at 137 min ⁻¹
29 Dec 2017 12:54	NonSust at 194 min ⁻¹
29 Dec 2017 12:53	NonSust at 152 min ⁻¹
29 Dec 2017 12:53	NonSust at 157 min ⁻¹
14 Dec 2017 11:09	Tachy at 177 min ⁻¹
14 Dec 2017 11:08	Tachy at 181 min ⁻¹
14 Dec 2017 11:07	Tachy at 186 min ⁻¹
14 Dec 2017 11:06	Tachy at 180 min ⁻¹
14 Dec 2017 11:05	Tachy at 166 min ⁻¹
14 Dec 2017 11:03	Tachy at 186 min ⁻¹
14 Dec 2017 11:02	Tachy at 174 min ⁻¹
14 Dec 2017 11:02	Tachy at 147 min⁻¹
14 Dec 2017 10:55	Tachy at 182 min ⁻¹
14 Dec 2017 10:49	Tachy at 168 min ⁻¹
13 Dec 2017 08:12	Tachy at 174 min ⁻¹
30 Nov 2017 15:03	Tachy at 168 min ⁻¹
30 Nov 2017 15:02	Tachy at 180 min ⁻¹
30 Nov 2017 15:01	Tachy at 166 min ⁻¹

2869 Software Version: 2.02

J275 Firmware Version: D_v1.02.00(1.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 4 of 5

ſ	Clinician Signature:
l	
l	

All Events Since Last Reset	(Continued)	(02 Jun 2014)
------------------------------------	-------------	---------------

Events office East Neset	
30 Nov 2017 14:59	Tachy at 172 min ⁻¹
30 Nov 2017 14:56	Tachy at 172 min⁻¹
30 Nov 2017 14:52	Tachy at 172 min ⁻¹
30 Nov 2017 14:48	Tachy at 175 min ⁻¹
30 Nov 2017 14:42	Tachy at 170 min ⁻¹
28 Nov 2017 18:22	Tachy at 160 min ⁻¹
13 Nov 2017 08:01	Tachy at 175 min ⁻¹
12 Nov 2017 09:39	Tachy at 172 min ⁻¹
12 Nov 2017 09:38	Tachy at 139 min ⁻¹
12 Nov 2017 09:36	Tachy at 175 min ⁻¹
12 Nov 2017 09:34	Tachy at 173 min ⁻¹
09 Nov 2017 22:34	Tachy at 166 min⁻¹
06 Nov 2017 20:05	Tachy at 175 min ⁻¹
06 Nov 2017 14:57	Tachy at 165 min ⁻¹
06 Nov 2017 14:18	Tachy at 165 min⁻¹
06 Nov 2017 11:13	Tachy at 166 min ⁻¹
06 Nov 2017 11:06	Tachy at 173 min⁻¹
12 Oct 2017 06:31	Tachy at 175 min ⁻¹
12 Oct 2017 06:23	Tachy at 172 min ⁻¹
12 Oct 2017 06:16	Tachy at 166 min⁻¹
12 Oct 2017 05:49	Tachy at 169 min⁻¹
12 Oct 2017 05:48	Tachy at 170 min⁻¹
12 Oct 2017 05:38	Tachy at 167 min⁻¹
12 Oct 2017 05:30	Tachy at 148 min⁻¹
12 Oct 2017 05:28	Tachy at 179 min⁻¹
12 Oct 2017 05:25	Tachy at 163 min ⁻¹
12 Oct 2017 05:23	Tachy at 181 min ⁻¹
12 Oct 2017 05:22	Tachy at 163 min ⁻¹
10 Oct 2017 22:47	Tachy at 174 min⁻¹
14 Sep 2017 10:27	Tachy at 164 min ⁻¹
12 Sep 2017 18:06	Tachy at 182 min⁻¹
24 Aug 2017 12:25	Tachy at 169 min⁻¹
24 Aug 2017 12:14	Tachy at 169 min⁻¹
24 Aug 2017 12:12	Tachy at 171 min⁻¹
24 Aug 2017 11:59	Tachy at 143 min⁻¹
23 Jul 2017 08:31	Tachy at 167 min ⁻¹
26 Jan 2015 13:34	RV Auto