

ZOOM ® View™ Report Created 05 Aug 2020

**QUICK NOTES ® Report** 

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L311/269927

Last Office Interrogation

03 Apr 2019 Implant Date N/R

## My Alerts

There are no alerts to display.

## **Events Since Last Reset (23 Mar 2017)**

25 Jul 2020 12:06 RYTHMIQ at 45 min<sup>-1</sup>
25 Jul 2020 11:48 RYTHMIQ at 64 min<sup>-1</sup>
See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 6 years
Magnet Rate 100 min<sup>-1</sup>

One Year Remaining

Explant

	—		
Leads Data	Implant N/R N/R N/R	Previous Session	Most Recent
Atrial			
Intrinsic Amplitude	N/R mV	2.4 mV	3.2 mV
Pace Impedance	N/R $\Omega$	740 Ω	$765~\Omega$
Pace Threshold	N/R V @ N/R ms	Auto 0.9 V @ 0.4 ms	Test Failed
Ventricular			
Intrinsic Amplitude	N/R mV	15.4 mV	8.4 mV @ 38 min <sup>-1</sup>
Pace Impedance	N/R $\Omega$	660 Ω	632 Ω
Pace Threshold	N/R V @ N/R ms	Auto 0.5 V @ 0.4 ms	Auto 1.0 V @ 0.4 ms

**Brady Counters Since Last Reset (23 Mar 2017)** 

Counters

Atrial 86% Paced Ventricular 52% Paced

**Atrial Arrhythmia** 

% AT/AF 0

(One Year of Data)

**Settings** 

Ventricular Tachy Settings Atrial Tachy Settings

Ventricular Tachy EGM Storage On ATR Mode Switch 170 min<sup>-1</sup> DDI

Detection Rate 160 min<sup>-1</sup>

**Brady Settings** 

Mode DDDR Pacing Output

RYTHMIQ $^{\text{TM}}$  AAIR With Atrial Auto 2.0 V @ 0.4 ms VVI Backup Ventricular Trend 2.5 V @ 0.4 ms

Lower Rate Limit 55 min<sup>-1</sup> Sensitivity

Maximum Tracking Rate 130 min<sup>-1</sup> Atrial Fixed 0.5 mV Maximum Sensor Rate 130 min<sup>-1</sup> Ventricular Fixed 2.0 mV

Paced AV Delay 80 - 180 ms Leads Configuration (Pace/Sense)

Sensed AV Delay 65 - 150 ms Atrial Bipolar A-Refractory (PVARP) 240 - 280 ms Ventricular Bipolar V-Refractory (VRP) 230 - 250 ms

All Events Since Last Reset (23 Mar 2017)

04 Aug 2020 15:24 RA Auto 04 Aug 2020 15:23 RV Auto

25 Jul 2020 12:06 RYTHMIQ at 45 min<sup>-1</sup>

2869 Software Version: 2.06 L311 Firmware Version: F\_v1.00.00(1.10)

Boston Scientific Corporation or its affiliates. All rights reserved.

Page 1 of 2

© 2014

Clinician Signature:

## All Events Since Last Reset (Continued) (23 Mar 2017)

25 Jul 2020 11:50	PMT at 130 min <sup>-1</sup>
25 Jul 2020 11:48	RYTHMIQ at 64 min <sup>-1</sup>
25 Jul 2020 11:44	PMT at 130 min <sup>-1</sup>
25 Jul 2020 11:43	PMT at 130 min <sup>-1</sup>
25 Jul 2020 11:42	RYTHMIQ at 62 min <sup>-1</sup>
25 Jul 2020 11:20	PMT at 130 min <sup>-1</sup>
25 Jul 2020 11:16	PMT at 130 min <sup>-1</sup>
25 Jul 2020 11:03	RYTHMIQ at 63 min <sup>-1</sup>
17 Jul 2020 11:04	RYTHMIQ at 47 min <sup>-1</sup>
17 Jul 2020 10:40	RYTHMIQ at 46 min <sup>-1</sup>
16 Jul 2020 22:18	RYTHMIQ at 45 min <sup>-1</sup>
22 May 2020 15:09	RYTHMIQ at 56 min <sup>-1</sup>
11 May 2020 17:53	RYTHMIQ at 62 min <sup>-1</sup>
07 May 2020 02:14	RYTHMIQ at 44 min <sup>-1</sup>
05 May 2020 10:01	RYTHMIQ at 57 min <sup>-1</sup>