



ZOOM® View™

QUICK NOTES® Report

Report Created 02 Jul 2014

Date of Birth N/R N/R N/R
Device INVIVE W173/118030

Last Office Interrogation
N/R
Implant Date
N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (02 Jul 2014)

There are no events to display.

Battery OK

Approximate time to explant: > 7 years
Magnet Rate 100 min⁻¹

One Year Remaining

Explant



Leads Data

Implant
N/R N/R N/R

Previous Session

Most Recent

Atrial

Intrinsic Amplitude	N/R mV	N/R mV	N/R mV
Pace Impedance	N/R Ω	N/R Ω	>2000 Ω
Pace Threshold	N/R V @ N/R ms	N/R	

Right Ventricular

Intrinsic Amplitude	N/R mV	N/R mV	PVC mV @ N/R min ⁻¹
Pace Impedance	N/R Ω	N/R Ω	874 Ω
Pace Threshold	N/R V @ N/R ms	N/R	Auto 0.5 V @ 0.4 ms

Left Ventricular

Intrinsic Amplitude	N/R mV	N/R mV	PVC mV
Pace Impedance	N/R Ω	N/R Ω	1347 Ω
Pace Threshold	N/R V @ N/R ms	N/R	0.8 V @ 0.4 ms

Brady/CRT Counters Since Last Reset (02 Jul 2014)

Atrial	0% Paced
Right Ventricular	4% Paced
Left Ventricular	3% Paced

Settings

Ventricular Tachy Settings

Ventricular Tachy EGM Storage On
Detection Rate 160 min⁻¹

Brady/CRT Settings

Mode	VVIR	Pacing Output	
Lower Rate Limit	80 min ⁻¹	Right Ventricular	Auto 2.0 V @ 0.4 ms
Maximum Sensor Rate	130 min ⁻¹	Left Ventricular	3.0 V @ 0.4 ms
RV-Refractory (RVRP)	230 - 250 ms	Sensitivity	
LV-Refractory (LVRP)	250 ms	Atrial	Fixed 0.75 mV
Ventricular Pacing Chamber	BiV	Right Ventricular	Fixed 2.5 mV
LV Offset	-40 ms	Left Ventricular	Fixed 2.5 mV
		Leads Configuration (Pace/Sense)	
		Atrial	--/Bipolar
		Right Ventricular	Bipolar
		Left Ventricular	LVtip>>LVring