Scientific

ZOOM ® View™		Report Created 11 Dec 2020
QUICK NOTES ® Report		Last Office Interrogation
		10-JUL-2013
		Implant Date
Device	ALTRUA 50 S508/576134	N.R. N.R. N.R.

Events Since Last Reset

09-JUL-2013 12:21 Tach at 190 min⁻¹
09-JUL-2013 12:14 Tach at 280 min⁻¹
See last page for full list of Events Since Last Reset.

Battery Good

Longevity Remaining 3.5 years

@ Current Pacing Percentage(s)

Magnet Rate 100 min⁻¹

Elective Replacement (ERT)
End of Life (EOL)

Beginning of Life (BOL)

Previous Session Leads Data Implant **Present Session** N.R. N.R. N.R. 10-JUL-2013 Intrinsic Amplitude N.R. mV 5 mV 3.3 mV Pace Impedance N.R. Ω 580 Ω 560 Ω Pace Threshold N.R. V 1.4 V @ 0.6 ms

Brady Counters Since Last Reset

86 % Paced

Settings
Tachy Settings

Tachy Detection On Rate 170 min⁻¹

Duration 8 cycles

Brady Settings Pacing Output 5.0 V @ 0.60 ms Mode SSI Sensitivity 1.0 mV Lower Rate Limit 80 min⁻¹ Lead Configuration **Bipolar** Maximum Tracking Rate -- min⁻¹ Pace Refractory 250 ms Sense

All Events Since Last Reset

09-JUL-2013 12:21 Tach at 190 min⁻¹
09-JUL-2013 12:14 Tach at 280 min⁻¹
09-JUL-2013 12:09 Tach at 240 min⁻¹
09-JUL-2013 12:09 Tach at 260 min⁻¹
09-JUL-2013 12:07 Tach at 220 min⁻¹

2892 Software Version: 1.00 S508 Firmware Version: 0a © 2010

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature: