



ZOOM ® View™

## QUICK NOTES ® Report

Report Created 04 Jun 2014

Date of Birth N/R N/R N/R  
Device VITALIO MRI J275/101893

Last Office Interrogation  
N/R  
Implant Date  
N/R

## My Alerts

There are no alerts to display.

## Events Since Last Reset (02 Jun 2014)

04 Jun 2014 04:17 NonSust at 305 min<sup>-1</sup>

## Battery OK

Approximate time to explant: > 8 years  
Magnet Rate 100 min<sup>-1</sup>

One Year Remaining

Explant



## Leads Data

## Implant

## Previous Session

## Most Recent

	N/R N/R N/R		
Intrinsic Amplitude	N/R mV	N/R mV	24.9 mV @ 48 min <sup>-1</sup>
Pace Impedance	N/R Ω	N/R Ω	616 Ω
Pace Threshold	N/R V @ N/R ms	N/R	Auto 0.6 V @ 0.4 ms

## Brady Counters Since Last Reset (02 Jun 2014)

% Paced 74

## Settings

## Tachy Settings

Tachy EGM Storage On  
Detection Rate 160 min<sup>-1</sup>

## Brady Settings

Mode	VVIR	Pacing Output	
Lower Rate Limit	55 min <sup>-1</sup>		Auto 1.1 V @ 0.4 ms
Maximum Sensor Rate	115 min <sup>-1</sup>	Sensitivity	
Refractory	250 - 250 ms		Fixed 2.5 mV
		Leads Configuration (Pace/Sense)	Unipolar/Bipolar