

ZOOM ® View™	Report Created 17 Feb 2022
--------------	----------------------------

Combined Follow-up Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L310/836133

Last Office Interrogation

26 Aug 2021 Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (02 Jul 2021)

17 Feb 2022 02:51 RV Auto

Battery OK

Approximate time to explant: 10 years Magnet Rate 100 min⁻¹

One Year Remaining

Explant

Leads Data Implant Previous Session Most Recent N/R N/R N/R 13.8 mV @ 47 min⁻¹ Intrinsic Amplitude N/R mV 14.1 mV 701 Ω Pace Impedance $N/R \Omega$ 849 Ω Pace Threshold N/R V @ N/R ms 0.4 V @ 0.4 ms 0.7 V @ 0.4 ms **Settings**

Tachy Settings

Tachy EGM Storage On

Detection Rate 160 min⁻¹

Brady Settings

Mode VVIR Pacing Output

Lower Rate Limit 60 min⁻¹ Trend 2.5 V @ 0.4 ms

Maximum Sensor Rate 110 min⁻¹ Sensitivity

Refractory 250 - 280 ms Fixed 2.0 mV

Leads Configuration (Pace/Sense)

Bipolar

Rate Adaptive Pacing

Minute Ventilation On Accelerometer Passive

2869 Software Version: 2.06


L310 Firmware Version: F_v1.00.00(1.10)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 4

Clinician Signature:



2869 Software Version: 2.06 L310 Firmware Version: F_v1.00.00(1.10) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 4

Clinician Signature:

Tachy Counters	Since Last Reset 02 Jul 2021	Device Totals
Episode Counters		
Total Episodes	0	0
Tachy Episodes	0	0
Nonsustained Episodes	0	0
Brady Counters	Reset Before Last N/R day(s) N/R to 02 Jul 2021	Since Last Reset 230 day(s) 02 Jul 2021 to Today
Counters		·
% Paced	0	80
Intrinsic Promotion		
Rate Hysteresis		
% Successful	0	0

Histograms Reset Before Last (N/R)



Histograms Since Last Reset (02 Jul 2021)

