

ZOOM ® View[™] Report Created 18 Sep 2015

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device AUTOGEN X4 CRT-D G179/135761

Tachy Mode Monitor + Therapy

Last Office Interrogation

16 Sep 2015 Implant Date 16 Sep 2015

My Alerts

There are no alerts to display.

Events Since Last Reset (16 Sep 2015)

18 Sep 2015 08:50 ATR at 323 min⁻¹, Avg V Rate in ATR: 77 min⁻¹ 18 Sep 2015 08:50 ATR at 263 min⁻¹, Avg V Rate in ATR: 130 min⁻¹

See last page for full list of Events Since Last Reset.

D-44		ł
Batterv	UN	

 $\begin{array}{lll} \mbox{Approximate time to explant:} & > 5 \mbox{ years} \\ \mbox{Charge Time} & \mbox{N/R s} \\ \mbox{Last Capacitor Re-form} & \mbox{N/R} \\ \end{array}$



Leads Data	Implant 16 Sep 2015	Previous Session	Most Recent
Atrial			
Intrinsic Amplitude	N/R mV	0.5 mV	1.3 mV
Pace Impedance	N/R Ω	$605~\Omega$	$600~\Omega$
Pace Threshold	N/R V @ N/R ms	N/R	N/R: rate too high
Right Ventricular			_
Intrinsic Amplitude	N/R mV	9.1 mV	22.7 mV @ 118 min ⁻¹
Pace Impedance	N/R Ω	619 Ω	470 Ω
Pace Threshold	N/R V @ N/R ms	0.3 V @ 0.4 ms	0.3 V @ 0.4 ms
Left Ventricular			
Intrinsic Amplitude	N/R mV	17.7 mV	20.0 mV
Pace Impedance	N/R Ω	550 Ω	541 Ω
Pace Threshold	N/R V @ N/R ms	Auto 0.5 V @ 0.4 ms	1.4 V @ 0.4 ms
Shock Vector			
Shock Impedance	N/R Ω	73 Ω	64 Ω
Brady/CRT Counters Since Las	t Roset (16 Sen 2015)		

Brady/CRT Counters Since Last Reset (16 Sep 2015)

Atrial <1% Paced Right Ventricular 28% Paced Left Ventricular 86% Paced

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 41J, 41Jx6 VT 175 min⁻¹ Burst Ramp 41J, 41Jx4

VT-1 150 min⁻¹ Monitor Only

Brady/CRT Settings

DDIR	Pacing Output	
55 min⁻¹	Atrial	Trend 3.5 V @ 0.4 ms
130 min ⁻¹	Right Ventricular	Trend 3.5 V @ 0.4 ms
180 - 180 ms	Left Ventricular	Auto 1.7 V @ 0.4 ms
240 - 280 ms	Sensitivity	
230 - 250 ms	Atrial	AGC 0.25 mV
250 ms	Right Ventricular	AGC 0.6 mV
BiV	Left Ventricular	AGC 1.0 mV
-20 ms	Leads Configuration (Pa	ce/Sense)
	Atrial	Bipolar
	55 min ⁻¹ 130 min ⁻¹ 180 - 180 ms 240 - 280 ms 230 - 250 ms 250 ms BiV	55 min ⁻¹ Atrial 130 min ⁻¹ Right Ventricular 180 - 180 ms Left Ventricular 240 - 280 ms Sensitivity 230 - 250 ms Atrial 250 ms Right Ventricular BiV Left Ventricular -20 ms Leads Configuration (Pa

2868 Software Version: 3.05

G179 Firmware Version: E_v1.02.00(1.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 2

Clinician Signature:

Settings (Continued)

Brady/CRT Settings

Leads Configuration (Pace/Sense)

Right Ventricular LV (Pace) LV (Sense) Bipolar LVRing3>>Can LVTip1>>LVRing2

All Events Since Last Reset (16 Sep 2015)

18 Sep 2015 08:50	ATR at 323 min ⁻¹ , Avg V Rate in ATR: 77 min ⁻¹
18 Sep 2015 08:50	ATR at 263 min ⁻¹ , Avg V Rate in ATR: 130 min ⁻¹
18 Sep 2015 08:49	ATR at 317 min ⁻¹ , Avg V Rate in ATR: 128 min ⁻¹
18 Sep 2015 08:49	ATR at 305 min ⁻¹ , Avg V Rate in ATR: 83 min ⁻¹
18 Sep 2015 06:32	RV Auto
18 Sep 2015 03:33	LV Auto
16 Sep 2015 10:30	ATR at 323 min ⁻¹ , Avg V Rate in ATR: 76 min ⁻¹
16 Sep 2015 10:30	ATR at 309 min ⁻¹ , Avg V Rate in ATR: 73 min ⁻¹
16 Sep 2015 10:29	ATR at 264 min ⁻¹ , Avg V Rate in ATR: 73 min ⁻¹
16 Sep 2015 10:29	ATR at 260 min ⁻¹ . Avg V Rate in ATR: 90 min ⁻¹

Clinician Signature: