

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L311/269927

Last Office Interrogation

N/R Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (23 Mar 2017)

26 Mar 2017 19:10 RA Auto 26 Mar 2017 18:16 RV Auto

Battery OK

Approximate time to explant: > 8 years

Magnet Rate 100 min⁻¹



Implant Leads Data Most Recent Previous Session N/R N/R N/R **Atrial** Intrinsic Amplitude N/R mV N/R mV 2.2 mV Pace Impedance N/R Ω $N/R \Omega$ 615 Ω Pace Threshold Auto 1.7 V @ 0.4 ms N/R V @ N/R ms N/R Ventricular 14.8 mV @ 40 min⁻¹ Intrinsic Amplitude N/R mV N/R mV Pace Impedance $N/R \Omega$ $N/R \Omega$ 687 Ω Pace Threshold N/R V @ N/R ms N/R Auto 0.7 V @ 0.4 ms

Brady Counters Since Last Reset (23 Mar 2017)

Counters

Atrial 96% Paced Ventricular 100% Paced

Atrial Arrhythmia

% AT/AF 0

Settings

Ventricular Tachy Settings Atrial Tachy Settings

Ventricular Tachy EGM Storage On ATR Mode Switch 170 min⁻¹ DDI

Detection Rate 160 min⁻¹

Brady Settings

Mode DDDR Pacing Output

RYTHMIQTM Off Atrial Trend 3.0 V @ 0.4 msLower Rate Limit 55 min^{-1} Ventricular Trend 3.0 V @ 0.4 ms

Maximum Tracking Rate 130 min⁻¹ Sensitivity

Maximum Sensor Rate 130 min⁻¹ Atrial Fixed 0.75 mV Paced AV Delay 80 - 180 ms Ventricular Fixed 2.5 mV

Sensed AV Delay 65 - 150 ms Leads Configuration (Pace/Sense)

A-Refractory (PVARP) 240 - 280 ms Atrial Bipolar V-Refractory (VRP) 230 - 250 ms Ventricular Bipolar

2869 Software Version: 2.02 L311 Firmware Version: F_v1.00.00 © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature: