

	<b>ZOOM ® View™</b>		<b>Report Created 08 Jan 2018</b>
	<b>Device Settings Report</b>		Last Office Interrogation <b>09 Jan 2017</b>
	Date of Birth	N/R N/R N/R	Implant Date
	Device	VITALIO MRI J275/101893	<b>N/R</b>

### Programming

Last Programmed 08 Jan 2018

### Tachy

Tachy EGM Storage On  
Detection Rate 160 min<sup>-1</sup> (375 ms)

### Atrial Tachy

#### Therapy

#### Ventricular Regulation

Vent Rate Regulation Off

## Brady

### Settings

Mode	VVIR	Output	
Lower Rate Limit	55 min <sup>-1</sup>		Auto 1.5 V @ 0.4 ms
Maximum Sensor Rate	115 min <sup>-1</sup>	Sensitivity	
Refractory	250 - 250 ms		Fixed 2.0 mV
Magnet Response	Pace Async	Leads	
Noise Response	VOO	Pace	Unipolar
		Sense	Bipolar
<b>Rate Enhancements</b>		<b>Rate Adaptive Pacing</b>	
Rate Smoothing		Minute Ventilation	On
Up	Off %	Response Factor	8
Down	Off %	Fitness Level	Active
		Ventilatory Threshold	120 min <sup>-1</sup>
		Ventilatory Thresh. Response	70 %
		Accelerometer	Passive

## Setup

### Telemetry

Enable use of ZIP™ telemetry	On
------------------------------	----

### Trending Setup

Recording Method	30 Second Average
Duration	25 hours
Data Storage	Continuous

### Sleep Schedule

Sleep Start Time	23:00 hh:mm
Sleep Duration	07 hours

### Leads Status Setup

	Daily Intrinsic Amplitude	Daily Impedance	Impedance Limits Low High	Pace Threshold
Pace/Sense	On	On	200 - 2000 Ω	On