

ZOOM ® View™ Report Created 02 Oct 201	
--	--

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J277/433364

Last Office Interrogation

N/R Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (30 Sep 2015)

02 Oct 2015 06:13 ATR at 191 min⁻¹, Avg V Rate in ATR: 90 min⁻¹ 02 Oct 2015 04:50 ATR at 68 min⁻¹, Avg V Rate in ATR: 68 min⁻¹

See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: > 11 years
Magnet Rate 100 min⁻¹



	Explain -			
Leads Data	Implant N/R N/R N/R	Previous Session	Most Recent	
Atrial				
Intrinsic Amplitude	N/R mV	N/R mV	2.5 mV	
Pace Impedance	N/R Ω	N/R Ω	408 Ω	
Pace Threshold	N/R V @ N/R ms	N/R	Auto 0.6 V @ 0.4 ms	
Ventricular				
Intrinsic Amplitude	N/R mV	N/R mV	8.6 mV @ 46 min ⁻¹	
Pace Impedance	N/R Ω	N/R Ω	419 Ω	
Pace Threshold	N/R V @ N/R ms	N/R	Auto 1.0 V @ 0.4 ms	

Brady Counters Since Last Reset (30 Sep 2015)

Counters

Atrial 77% Paced Ventricular 94% Paced

Atrial Arrhythmia

% AT/AF <1

Settings

Ventricular Tachy Settings	Atrial Tachy Settings		
Ventricular Tachy EGM Storage	On	ATR Mode Switch 170 min ⁻¹ DDI	
Detection Rate	160 min ^{−1}		
Brady Settings			
Mode	DDDR	Pacing Output	

RYTHMIQ™	Off	Atrial	Trend 3.5 V @ 0.4 ms
Lower Rate Limit	55 min ^{−1}	Ventricular	Trend 3.5 V @ 0.4 ms
Maximum Tracking Rate	130 min ⁻¹	Sensitivity	
Maximum Sensor Rate	130 min ⁻¹	Atrial	Fixed 0.75 mV

Paced AV Delay 80 - 180 ms Ventricular Fixed 2.5 mV

Sensed AV Delay 65 - 150 ms Leads Configuration (Pace/Sense)

A-Refractory (PVARP) 240 - 280 ms Atrial Bipolar V-Refractory (VRP) 230 - 250 ms Ventricular Bipolar

All Events Since Last Reset (30 Sep 2015)

02 Oct 2015 06:13 ATR at 191 min⁻¹, Avg V Rate in ATR: 90 min⁻¹
02 Oct 2015 04:50 ATR at 68 min⁻¹, Avg V Rate in ATR: 68 min⁻¹
02 Oct 2015 04:49 ATR at 62 min⁻¹, Avg V Rate in ATR: 65 min⁻¹
02 Oct 2015 04:47 ATR at 333 min⁻¹, Avg V Rate in ATR: 80 min⁻¹
02 Oct 2015 04:42 ATR at 188 min⁻¹, Avg V Rate in ATR: 107 min⁻¹

2869 Software Version: 2.02

J277 Firmware Version: D_v1.02.00(1.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 2

Clinician Signature:

All Events Since Last Reset (Continued) (30 Sep 2015)

02 Oct 2015 04:41	ATR at 86 min ⁻¹ , Avg V Rate in ATR: 75 min ⁻¹
02 Oct 2015 04:41	ATR at 101 min ⁻¹ , Avg V Rate in ATR: 61 min ⁻¹
01 Oct 2015 17:34	ATR at 287 min ⁻¹ , Avg V Rate in ATR: 84 min ⁻¹
01 Oct 2015 11:57	ATR at 61 min ⁻¹ , Avg V Rate in ATR: 60 min ⁻¹
01 Oct 2015 10:08	ATR at 75 min ⁻¹ , Avg V Rate in ATR: 67 min ⁻¹

2869 Software Version: 2.02 J277 Firmware Version: D_v1.02.00(1.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: