

Combined Follow-up Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI L310/836133 Last Office Interrogation

26 Aug 2021 Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (02 Jul 2021)

OK

17 Feb 2022 02:51 **RV** Auto

Battery

Approximate time to explant: 10.5 years 100 min⁻¹ Magnet Rate

One Year Remaining Explant

Leads Data Implant Previous Session Most Recent N/R N/R N/R 13.8 mV @ 47 min⁻¹ Intrinsic Amplitude N/R mV 14.1 mV 701 Ω Pace Impedance $N/R \Omega$ 849 Ω Pace Threshold N/R V @ N/R ms 0.4 V @ 0.4 ms 0.7 V @ 0.4 ms

Settings Tachy Settings

Tachy EGM Storage

On **Detection Rate** 160 min⁻¹

Brady Settings

VVI Mode **Pacing Output**

Lower Rate Limit 60 min⁻¹ Trend 2.5 V @ 0.4 ms

250 ms Sensitivity Refractory

Fixed 2.0 mV

Leads Configuration (Pace/Sense)

Bipolar

Rate Adaptive Pacing


Minute Ventilation **Passive** Accelerometer **Passive**

2869 Software Version: 2.06

L310 Firmware Version: F_v1.00.00(1.10)

© 2014 Boston Scientific Corporation or its affiliates. All rights reserved. Page 1 of 4

Clinician Signature:



2869 Software Version: 2.06 L310 Firmware Version: F_v1.00.00(1.10) © 2014


Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 4


Clinician Signature:

Tachy Counters	Since Last Reset 02 Jul 2021	Device Totals	
Episode Counters			
Total Episodes	0	0	
Tachy Episodes	0	0	
Nonsustained Episodes	0	0	
Brady Counters	Reset Before Last N/R day(s) N/R to 02 Jul 2021	Since Last Reset 230 day(s) 02 Jul 2021 to Today	
Counters		·	
% Paced	0	80	
Intrinsic Promotion			
Rate Hysteresis			
% Successful	0	0	

Histograms Reset Before Last (N/R)



Histograms Since Last Reset (02 Jul 2021)



2869 Software Version: 2.06 L310 Firmware Version: F_v1.00.00(1.10) © 2014
Boston Scientific Corporation
or its affiliates. All rights reserved.
Page 4 of 4

Clinician Signature: