Boston Scientific	ZOOM ® View™	**Report Created 18 May 2020**
	Combined Follow-up Report	Last Office Interrogation 20-JAN-2020 Implant Date N.R. N.R. N.R.
	Device ALTRUA 50 S501/590330	

Events Since Last Reset (21-JAN-2019)

There are no events to display.

Battery   Good

Longevity Remaining @ Current Pacing Percentage(s)	4.0 years		
Magnet Rate	100 min^{-1}	Elective Replacement (ERT) End of Life (EOL)	Beginning of Life (BOL)

Leads Data

	Implant N.R. N.R. N.R.	Previous Session	Present Session 20-JAN-2020
Intrinsic Amplitude	N.R. mV	7.4 mV	N.R. mV
Pace Impedance	N.R. Ω	460 Ω	440 Ω
Pace Threshold	N.R. V		1.1 V @ 0.4 ms

Settings

Tachy Settings

Tachy Detection	On		
Rate	170 min^{-1}		
Duration	8 cycles		

Brady Settings

		Pacing Output	AUTO 1.6V @ 0.40 ms
Mode	SSIR	Sensitivity	AUTO 1.48 mV
Lower Rate Limit	55 min^{-1}	Lead Configuration	Bipolar
Maximum Tracking Rate	-- min^{-1}	Pace	--
Maximum Sensor Rate	130 min^{-1}	Sense	--
Refractory	250 ms		

Daily Measurement

Date	Amplitude (mV)	Impedance (Ω)	Threshold (V)
19-JAN-2020	N.R.	400	1.0
18-JAN-2020	>12.0	430	1.0
17-JAN-2020	>12.0	430	0.7
16-JAN-2020	N.R.	460	0.9
15-JAN-2020	N.R.	430	0.8
14-JAN-2020	>12.0	430	0.8
13-JAN-2020	>12.0	440	0.8
10-JAN-2020	>12.0	430	1.0
03-JAN-2020	>12.0	430	1.0
27-DEC-2019	>12.0	440	1.0
20-DEC-2019	>12.0	440	1.0
13-DEC-2019	>12.0	420	1.0
06-DEC-2019	11.7	430	1.0
29-NOV-2019	>12.0	420	1.0
22-NOV-2019	10.0	430	1.0
15-NOV-2019	6.7	430	1.0

Daily Measurement (Continued)

Date	Amplitude (mV)	Impedance (Ω)	Threshold (V)
08-NOV-2019	10.8	440	1.0
01-NOV-2019	6.2	430	1.0
25-OCT-2019	8.4	430	1.0
18-OCT-2019	6.5	420	1.1
11-OCT-2019	10.3	430	1.1
04-OCT-2019	8.6	440	1.0
27-SEP-2019	10.0	440	1.0
20-SEP-2019	7.3	430	1.0
13-SEP-2019	>12.0	440	1.0
06-SEP-2019	10.9	430	1.1
30-AUG-2019	9.0	430	1.0
23-AUG-2019	6.5	430	1.0
16-AUG-2019	8.8	430	1.1
09-AUG-2019	7.1	420	1.0
02-AUG-2019	8.7	420	1.1
26-JUL-2019	7.0	440	1.0
19-JUL-2019	9.1	420	1.0
12-JUL-2019	11.4	430	0.9
05-JUL-2019	11.1	430	1.0
28-JUN-2019	8.1	430	1.0
21-JUN-2019	>12.0	440	1.0
14-JUN-2019	10.5	430	1.0
07-JUN-2019	6.3	430	1.1
31-MAY-2019	10.8	430	1.1
24-MAY-2019	8.2	430	1.0
17-MAY-2019	8.6	430	1.0
10-MAY-2019	9.7	430	1.1
03-MAY-2019	9.1	430	1.1
26-APR-2019	>12.0	430	1.1
19-APR-2019	9.3	430	1.0
12-APR-2019	>12.0	440	1.1
05-APR-2019	10.4	430	1.1
29-MAR-2019	8.8	440	1.0
22-MAR-2019	>12.0	430	1.1
15-MAR-2019	7.8	440	1.1
08-MAR-2019	>12.0	440	1.1
01-MAR-2019	10.3	440	1.1
22-FEB-2019	>12.0	440	1.0
15-FEB-2019	8.1	440	1.0
08-FEB-2019	10.4	440	1.0
01-FEB-2019	6.7	440	1.1
25-JAN-2019	7.0	440	1.0

Activity Log

Date	Percent Active	Exercise Intensity
19-JAN-2020	1	3
18-JAN-2020	1	3
17-JAN-2020	1	2
16-JAN-2020	1	2
15-JAN-2020	1	2
14-JAN-2020	1	2
13-JAN-2020	1	2
11-JAN-2020	1	3
04-JAN-2020	1	3
28-DEC-2019	1	3
21-DEC-2019	1	4
14-DEC-2019	1	2
07-DEC-2019	1	3
30-NOV-2019	1	3
23-NOV-2019	1	3
16-NOV-2019	1	3
09-NOV-2019	1	4
02-NOV-2019	1	3
26-OCT-2019	1	3
19-OCT-2019	1	3
12-OCT-2019	1	3
05-OCT-2019	1	3
28-SEP-2019	1	3
21-SEP-2019	1	3
14-SEP-2019	1	3
07-SEP-2019	1	5
31-AUG-2019	1	3
24-AUG-2019	1	2
17-AUG-2019	1	2
10-AUG-2019	1	3
03-AUG-2019	1	3
27-JUL-2019	1	3
20-JUL-2019	1	2
13-JUL-2019	1	4
06-JUL-2019	1	3
29-JUN-2019	1	3
22-JUN-2019	1	3
15-JUN-2019	1	4
08-JUN-2019	1	3
01-JUN-2019	1	3
25-MAY-2019	1	4
18-MAY-2019	1	4
11-MAY-2019	1	3
04-MAY-2019	1	3
27-APR-2019	1	3

Activity Log (Continued)

Date	Percent Active	Exercise Intensity
20-APR-2019	1	3
13-APR-2019	1	3
06-APR-2019	1	3
30-MAR-2019	1	4
23-MAR-2019	1	3
16-MAR-2019	1	4
09-MAR-2019	1	4
02-MAR-2019	1	3
23-FEB-2019	1	3
16-FEB-2019	1	3
09-FEB-2019	1	3
02-FEB-2019	1	3
26-JAN-2019	1	3
19-JAN-2019	1	4

Trends

Events

| Feb 19	Mar 19	Apr 19	May 19	Jun 19	Jul 19	Aug 19	Sep 19	Oct 19	Nov 19	Dec 19	Jan 20

Percent Active

Exercise Intensity

Amplitude

Trends (Continued)

Impedance

Threshold

Tachy Counters		Since Last Reset 21-JAN-2019
Episode Counters		
Tachy Detections		0

Brady Counters	Percentages	Since Last Reset 21-JAN-2019
Counters		
Paced	95 %	28.4M
Sensed	5 %	1.4M
Intrinsic Promotion		
Rate Hysteresis		
Searches		0
Successful Searches		0
Counters		
Interval Variation		
Variation 0 <= 10%		27.5M
Variation 10 <= 20%		206.2K
Variation 20 <= 30%		188.7K
Variation > 30%		1.8M

Histograms
Since Last Reset(21-JAN-2019)

2892 Software Version: 1.00
S501 Firmware Version: 0a

© 2010
Boston Scientific Corporation
or its affiliates. All rights reserved.
Page 5 of 5

Clinician Signature: