

ZOOM ® View™ Report Created 08 Sep 2014

QUICK NOTES ® Report

Hlucík, Stepán

Date of Birth

Device

7 Jan 1944 INCEPTA CRT-D P162/119874

Tachy Mode Monitor + Therapy

Last Office Interrogation

2 Dec 2013 Implant Date 12 Mar 2013

My Alerts

a 08 Sep 2014 13:28 Check Atrial and LV Leads

Events Since Last Reset (14 Mar 2013)

VT at 200 min⁻¹, ATPx4, 17J 30 Aug 2014 13:14 VT at 181 min⁻¹, **♦** ATPx4, 17J 09 Aug 2014 16:46

See last page for full list of Events Since Last Reset.

Battery OK

> Approximate time to explant: 6 years Charge Time 9.6 s

> Last Capacitor Re-form 07 Sep 2014 07:19

One Year Remaining **Explant**

Leads Data Implant Previous Session Most Recent 12 Mar 2013 Atrial N/R mV Intrinsic Amplitude N/R mV N/R mV Pace Impedance $N/R \Omega$ >2000 Ω >2000 Ω N/R V @ N/R ms Pace Threshold 4.0 V @ 0.4 ms **Right Ventricular** Intrinsic Amplitude 7.0 mV PVC mV PVC mV @ N/R min⁻¹ Pace Impedance 650Ω 515 Ω 522 Ω Pace Threshold 0.5 V @ 0.4 ms 0.6 V @ 0.4 ms 1.0 V @ 0.4 ms Left Ventricular PVC mV PVC mV Intrinsic Amplitude 12.0 mV Pace Impedance 650Ω 539Ω 508Ω Pace Threshold 1.3 V @ 0.4 ms 1.4 V @ 0.4 ms 1.4 V @ 0.4 ms **Shock Vector** 45 Ω 61 Ω 57 Ω Shock Impedance

Brady/CRT Counters Since Last Reset (14 Mar 2013)

Atrial 0% Paced Right Ventricular 98% Paced Left Ventricular 99% Paced

Settings

Ventricular Tachy Settings

VF 230 min⁻¹ ATP 31J, 41J, 41Jx6 205 min⁻¹ Scan Ramp 31J, 41J, 41Jx4 VT-1 160 min⁻¹ Scan 41J, 41J, 41Jx3 Ramp

Brady/CRT Settings

Mode **VVIR** Pacing Output Lower Rate Limit 65 min⁻¹ Right Ventricular 3.5 V @ 0.4 ms 4.0 V @ 0.4 ms Maximum Sensor Rate 130 min⁻¹ Left Ventricular RV-Refractory (RVRP) 230 - 250 ms Sensitivity 250 ms AGC 0.25 mV LV-Refractory (LVRP) Atrial BiV Right Ventricular Ventricular Pacing Chamber AGC 0.6 mV LV Offset 0 ms Left Ventricular AGC 1.0 mV

Leads Configuration (Pace/Sense)

Bipolar/Off Atrial Right Ventricular **Bipolar**

2868 Software Version: 3.04

P162 Firmware Version: B_v1.02.00(3.01)

© 2010 **Boston Scientific Corporation** or its affiliates. All rights reserved. Page 1 of 2

Clinician Signature:

Settings (Continued)

Brady/CRT Settings

16 May 2013 17:54

17 Apr 2013 15:14

All Events Since Last Reset (14 Mar 2013)

Leads Configuration (Pace/Sense)

Left Ventricular LVring>>RV / LVtip>>LVring

| 30 Aug 2014 13:14 | VT at 200 min⁻¹, � ATPx4, 17J |
|-------------------|--------------------------------------------------|
| 30 Aug 2014 13:13 | NonSustV at 201 min ⁻¹ , Nonsustained |
| 30 Aug 2014 13:13 | NonSustV at 186 min ⁻¹ , Nonsustained |
| 09 Aug 2014 16:46 | VT at 181 min⁻¹, � ATPx4, 17J |
| 09 Aug 2014 16:45 | NonSustV at 186 min ⁻¹ , Nonsustained |
| 09 Aug 2014 16:45 | NonSustV at 164 min ⁻¹ , Nonsustained |
| 09 Aug 2014 16:45 | NonSustV at 175 min ⁻¹ , Nonsustained |
| 09 Aug 2014 16:38 | NonSustV at 170 min ⁻¹ , Nonsustained |
| 09 Aug 2014 16:38 | NonSustV at 169 min ⁻¹ , Nonsustained |
| 09 Aug 2014 14:40 | NonSustV at 178 min ⁻¹ , Nonsustained |
| 09 Aug 2014 14:40 | NonSustV at 181 min ⁻¹ , Nonsustained |
| 09 Aug 2014 14:40 | NonSustV at 180 min ⁻¹ , Nonsustained |

VT at 178 min⁻¹, **♦** ATPx2

VT at 179 min⁻¹, **♦** ATPx4, 17J

2868 Software Version: 3.04

P162 Firmware Version: B_v1.02.00(3.01)

© 2010

Boston Scientific Corporation or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: