

ZOOM ® View™ F	Report Created 22 Apr 2020
----------------	----------------------------

QUICK NOTES ® Report

Date of Birth N/R N/R N/R

Device ACCOLADE MRI EL L331/876849

Last Office Interrogation

N/R Implant Date N/R

My Alerts

There are no alerts to display.

Events Since Last Reset (20 Apr 2020)

21 Apr 2020 15:10 RV Auto 21 Apr 2020 14:10 RA Auto

Battery OK

Approximate time to explant: > 11 years
Magnet Rate 100 min⁻¹

One Year Remaining Explant



Leads Data Implant Previous Session Most Recent N/R N/R N/R **Atrial** Intrinsic Amplitude N/R mV N/R mV 8.7 mV Pace Impedance $N/R \Omega$ $N/R \Omega$ 708 Ω Pace Threshold N/R V @ N/R ms Auto 0.4 V @ 0.4 ms N/R Ventricular Intrinsic Amplitude N/R mV N/R mV 19.6 mV @ 155 min⁻¹ Pace Impedance 890 Ω $N/R \Omega$ $N/R \Omega$ Pace Threshold N/R V @ N/R ms N/R Auto 0.6 V @ 0.4 ms POST completed. Test started at 20 Apr 2020 09:55

Brady Counters Since Last Reset (20 Apr 2020)

Counters

Atrial 4% Paced Ventricular 30% Paced

Atrial Arrhythmia

% AT/AF 0

Settings

Ventricular Tachy Settings		Atrial Tachy Settings	
Ventricular Tachy EGM Storage	On	ATR Mode Switch 170 min ⁻¹	DDI
Detection Rate	160 min ⁻¹		
Brady Settings			
Mode	DDD	Pacing Output	
RYTHMIQ™	Off	Atrial	Trend 3.5 V @ 0.4 ms
Lower Rate Limit	55 min ^{−1}	Ventricular	Trend 3.5 V @ 0.4 ms
Maximum Tracking Rate	130 min ⁻¹	Sensitivity	
Paced AV Delay	80 - 220 ms	Atrial	Fixed 0.75 mV
Sensed AV Delay	75 - 200 ms	Ventricular	Fixed 2.5 mV
A-Refractory (PVARP)	240 - 280 ms	Leads Configuration (Pace/Sen	ise)
V-Refractory (VRP)	230 - 250 ms	Atrial	Bipolar

Ventricular

2869 Software Version: 2.06

L331 Firmware Version: F_v1.00.00(1.10)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 1

Clinician Signature:

Bipolar