



Marker Legend Report

Date of Birth N/R N/R N/R
Device ACCOLADE MRI L310/ 839486

Last Office Interrogation
N/R
Implant Date
N/R

AS	Atrial Sense - After Refractory
AS-Hy	Atrial Sense - In Hysteresis Offset
[AS]	Atrial Sense - During Blanking
AP	Atrial Pace - Lower Rate
AP↓	Atrial Pace - Rate Smoothing Down
AP↑	Atrial Pace - Rate Smoothing Up
AP-Hy	Atrial Pace - At Hysteresis Rate
AP-Sr	Atrial Pace - Sensor Rate
AP-Ns	Atrial Pace - Noise (asynchronous pacing)
AF	Atrial Tachy Zone Sense
VS	Ventricular Sense - After Refractory
[VS]	Ventricular Sense - Noise First Trigger
VS-Hy	Ventricular Sense - At Hysteresis Rate
VP	Ventricular Pace - Lower Rate
VP↓	Ventricular Pace - Rate Smoothing Down
VP↑	Ventricular Pace - Rate Smoothing Up
VP-BP	Ventricular Pace - Backup Pace
VP-Hy	Ventricular Pace - At Hysteresis Rate
VP-Sr	Ventricular Pace - Sensor Rate
VP-Ns	Ventricular Pace - Noise (asynchronous pacing)
VP-VR	Ventricular Pace - Ventricular Rate Regulation
VT	VT Zone Sense
AN	Atrial Rate Noise
VN	Ventricular Rate Noise
Epsd	Tachy Start Episode
EpsdEnd	Tachy End Episode
Fusion	Fusion Beat
LOC	Loss of Capture
--	Unclassified Event
##.#V	Amplitude Threshold Test
##.#mV	Intrinsic Amplitude Test
###.##ms	Pulse Width Threshold Test
####	Interval in ms
####Ω	Lead Impedance Test