

Combined Follow-up Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J275/101893

Last Office Interrogation

08 Jan 2018 Implant Date **2 Jun 2014**

My Alerts

There are no alerts to display.

Events Since Last Reset (02 Jun 2014)

02 Jan 2019 09:31 Tachy at 175 min⁻¹
02 Jan 2019 09:23 Tachy at 163 min⁻¹
See last page for full list of Events Since Last Reset.

Battery OK

Approximate time to explant: 6.5 years Magnet Rate 100 min⁻¹

One Year Remaining Explant

Leads Data	Implant 2 Jun 2014	Previous Session	Most Recent
Intrinsic Amplitude	N/R mV	20.5 mV	23.1 mV @ 53 min ⁻¹
Pace Impedance	N/R Ω	576 Ω	627 Ω
Pace Threshold	N/R V @ N/R ms	Auto 1.0 V @ 0.4 ms	Auto 0.8 V @ 0.4 ms
0.441			

Settings

Tachy Settings

Tachy EGM Storage On
Detection Rate 160 min⁻¹

Brady Settings

Mode VVIR Pacing Output

Lower Rate Limit 55 min⁻¹ Auto 1.3 V @ 0.4 ms

Maximum Sensor Rate 115 min⁻¹ Sensitivity

Refractory 250 - 250 ms Fixed 2.0 mV

Leads Configuration (Pace/Sense)

Unipolar/Bipolar

Rate Adaptive Pacing


Minute Ventilation On Accelerometer Passive

2869 Software Version: 2.02 J275 Firmware Version: D_v1.02.00(1.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.


Page 1 of 5

Clinician Signature:




Tachy Counters	Since Last Reset 02 Jun 2014	Device Totals	
Episode Counters			
Total Episodes	65.8K	65.8K	
Tachy Episodes	1.7K	1.7K	
Nonsustained Episodes	64.1K	64.1K	
Brady Counters	Reset Before Last N/R day(s) N/R to 02 Jun 2014	Since Last Reset 1680 day(s) 02 Jun 2014 to Today	
Counters			
% Paced	0	53	
Intrinsic Promotion			
Rate Hysteresis			
% Successful	0	0	

Histograms Reset Before Last (N/R)



Histograms Since Last Reset (02 Jun 2014)



All Events Since Last Reset (02 Jun 2014)

LVCIII3 OIIICC Last Nesct	(02 0dil 2017)
07 Jan 2019 08:38	RV Auto
02 Jan 2019 16:37	NonSust at 170 min ⁻¹
02 Jan 2019 16:37	NonSust at 143 min ⁻¹
02 Jan 2019 16:36	NonSust at 146 min ⁻¹
02 Jan 2019 16:36	NonSust at 172 min ⁻¹
02 Jan 2019 10:02	NonSust at 151 min ⁻¹
02 Jan 2019 09:46	NonSust at 140 min ⁻¹
02 Jan 2019 09:31	Tachy at 175 min ⁻¹
02 Jan 2019 09:30	NonSust at 146 min ⁻¹
02 Jan 2019 09:29	NonSust at 139 min ⁻¹
02 Jan 2019 09:26	NonSust at 171 min ⁻¹
02 Jan 2019 09:25	NonSust at 169 min ⁻¹
02 Jan 2019 09:23	Tachy at 163 min ⁻¹
02 Jan 2019 09:21	Tachy at 145 min ⁻¹
02 Jan 2019 09:21	Tachy at 151 min ⁻¹
02 Jan 2019 09:18	Tachy at 175 min ⁻¹
17 Dec 2018 10:10	Tachy at 179 min ⁻¹
17 Dec 2018 10:07	Tachy at 171 min ⁻¹
20 Nov 2018 15:21	Tachy at 164 min ⁻¹
20 Nov 2018 15:14	Tachy at 168 min⁻¹
20 Nov 2018 15:13	Tachy at 167 min ⁻¹
29 Oct 2018 09:21	Tachy at 172 min ⁻¹
08 Sep 2018 08:52	Tachy at 163 min ⁻¹
08 Sep 2018 08:46	Tachy at 170 min ⁻¹

2869 Software Version: 2.02

J275 Firmware Version: D_v1.02.00(1.01)

© 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 4 of 5

Clinician Signature:	

LVCIIIS OIIIOC LUST INCSCI	(Solitiliaca) (SE Sali
01 Sep 2018 09:51	Tachy at 162 min ⁻¹
01 Sep 2018 09:37	Tachy at 161 min ⁻¹
01 Sep 2018 09:36	Tachy at 161 min ⁻¹
30 Aug 2018 09:10	Tachy at 171 min⁻¹
30 Aug 2018 08:07	Tachy at 160 min ⁻¹
30 Aug 2018 08:01	Tachy at 160 min ⁻¹
06 Jul 2018 08:25	Tachy at 174 min⁻¹
10 May 2018 13:16	Tachy at 161 min ⁻¹
10 May 2018 13:15	Tachy at 160 min ⁻¹
10 May 2018 11:53	Tachy at 167 min ⁻¹
07 May 2018 08:39	Tachy at 165 min ⁻¹
07 May 2018 07:29	Tachy at 168 min ⁻¹
07 May 2018 06:24	Tachy at 162 min ⁻¹
07 May 2018 06:17	Tachy at 175 min ⁻¹
02 May 2018 19:51	Tachy at 160 min ⁻¹
02 May 2018 19:47	Tachy at 160 min ⁻¹
02 May 2018 19:46	Tachy at 160 min ⁻¹
28 Mar 2018 07:49	Tachy at 161 min ⁻¹
04 Mar 2018 09:29	Tachy at 174 min⁻¹
04 Mar 2018 09:24	Tachy at 170 min ⁻¹
04 Mar 2018 09:19	Tachy at 137 min ⁻¹
04 Mar 2018 09:17	Tachy at 174 min⁻¹
28 Feb 2018 09:17	Tachy at 173 min⁻¹
28 Feb 2018 09:11	Tachy at 176 min⁻¹
28 Feb 2018 08:58	Tachy at 171 min⁻¹
28 Feb 2018 08:57	Tachy at 169 min ⁻¹
28 Feb 2018 08:32	Tachy at 172 min⁻¹
28 Feb 2018 07:55	Tachy at 171 min⁻¹
28 Feb 2018 07:54	Tachy at 152 min⁻¹
28 Feb 2018 07:29	Tachy at 187 min⁻¹
17 Feb 2018 10:32	Tachy at 179 min⁻¹
09 Jan 2018 14:51	Tachy at 179 min⁻¹
09 Jan 2018 14:49	Tachy at 147 min ⁻¹
09 Jan 2018 14:49	Tachy at 180 min⁻¹
09 Jan 2018 14:45	Tachy at 169 min ⁻¹
09 Jan 2018 14:45	Tachy at 175 min ⁻¹
09 Jan 2018 14:41	Tachy at 168 min⁻¹