

ZOOM ® View™	Report Created 25 Nov 2016

Device Settings Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J276/104135

Last Office Interrogation

03 Dec 2015 Implant Date N/R

Programming

Last Programmed 25 Nov 2016

Ventricular Tachy

2869 Software Version: 2.02

J276 Firmware Version: D_v1.02.00(1.01)

Ventricular Tachy EGM Storage On

Detection Rate 160 min⁻¹ (375 ms)

Atrial Tachy

Therapy

ATR Mode Switch Details Atrial Tachy Response

ATR Mode Switch On Atrial Flutter Response On Trigger Rate 160 min⁻¹ Trigger Rate 160 min⁻¹ Duration 8 cycles PMT Termination On

Entry Count 8 cycles Atrial Overdrive

Exit Count 8 cycles APP/ProACt Off Fallback

Mode DDI

Time 00:30 mm:ss
ATR Fallback LRL 70 min⁻¹
Ventricular Rate Regulation On
Maximum Pacing Rate 130 min⁻¹

© 2014 Clinician Signature:

Boston Scientific Corporation or its affiliates. All rights reserved.

Page 1 of 2

Brady				
Settings				
Mode	DDDR	Output		
RYTHMIQ™	AAIR With VVI	●A	Auto 2.0 V @ 0.4 ms	
	Backup	■V	4.0 V @ 0.5 ms	
Lower Rate Limit	60 min ^{−1}	Sensitivity		
Maximum Tracking Rate	130 min ⁻¹	●A	Fixed 0.5 mV	
Maximum Sensor Rate	130 min ⁻¹	■V	Fixed 2.0 mV	
Paced AV Delay	120 - 220 ms	Leads		
Sensed AV Delay	80 - 150 ms	●A		
A-Refractory (PVARP)	240 - 280 ms	Pace	Bipolar	
V-Refractory (VRP)	230 - 250 ms	Sense	Bipolar	
PVARP after PVC	400 ms	Safety Switch	On	
AV Search +	On	■V		
Search AV Delay	300 ms	Pace	Bipolar	
Search Interval	32 cycles	Sense	Bipolar	
Blanking		Safety Switch	On	
A-Blank after V-Pace	125 ms	Rate Adaptive Pacing		
A-Blank after V-Sense	45 ms	Minute Ventilation	On	
V-Blank after A-Pace	65 ms	Response Factor	8	
Magnet Response	Pace Async	Fitness Level	Active	
Noise Response	DOO	Ventilatory Threshold	120 min ⁻¹	
Rate Enhancements		Ventilatory Thresh. Response	70 %	
Rate Smoothing		Accelerometer	Passive	
Up	Off %			
Down	Off %			
Sudden Brady Response	Off			
Setup				
Telemetry				
Enable use of ZIP™ telemetry	On			
Trending Setup				
Recording Method	30 Second Average			
Duration		25 hours		
Data Storage		Continuous		
Sleep Schedule				
Olassa Otant Times		00.00 hh.m.m		

Sleep Start Time Sleep Duration 23:00 hh:mm 07 hours

Impedance Limits

Leads Status Setup

		Intrinsic Amplitude	Impedance	Low	High	
●A	Pace/Sense	On	On	200 -	2000Ω	On
■ V	Pace/Sense	On	On	200 -	2000 Ω	Off

Daily

Daily

2869 Software Version: 2.02 J276 Firmware Version: D_v1.02.00(1.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature:

Pace Threshold