

ZOOM ® View™	Report Created 15 Jan 2018

Device Settings Report

Date of Birth N/R N/R N/R

Device VITALIO MRI J276/102818

Last Office Interrogation

17 Jan 2017 Implant Date N/R

Programming

Last Programmed 19 Sep 2016

Ventricular Tachy

Ventricular Tachy EGM Storage On

Detection Rate 160 min⁻¹ (375 ms)

Atrial Tachy

Therapy

ATR Mode Switch Details Atrial Tachy Response

ATR Mode Switch
On
Atrial Flutter Response
On
Trigger Rate
170 min⁻¹
Trigger Rate
170 min⁻¹
Duration
8 cycles
PMT Termination
On

Entry Count 8 cycles Atrial Overdrive

Exit Count 8 cycles APP/ProACt Off

Fallback

Mode DDI

Time 00:30 mm:ss
ATR Fallback LRL 70 min⁻¹
Ventricular Rate Regulation On
Maximum Pacing Rate 130 min⁻¹

2869 Software Version: 2.02 J276 Firmware Version: D_v1.02.00(1.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 1 of 2

Clinician Signature:

Drody			
Brady			
Settings	DDDD	Outrost	
Mode	DDDR	Output	A 1- 0.03/ @ 0.4
RYTHMIQ™	AAIR With VVI	●A ■N	Auto 2.0 V @ 0.4 ms
D 4 11 11	Backup	■V	2.3 V @ 0.4 ms
Lower Rate Limit	55 min ⁻¹	Sensitivity	
Maximum Tracking Rate	130 min ⁻¹	●A	Fixed 0.5 mV
Maximum Sensor Rate	130 min ⁻¹	■ V	Fixed 2.0 mV
Paced AV Delay	80 - 180 ms	Leads	
Sensed AV Delay	65 - 150 ms	●A	
A-Refractory (PVARP)	240 - 280 ms	Pace	Bipolar
V-Refractory (VRP)	230 - 250 ms	Sense	Bipolar
PVARP after PVC	400 ms	Safety Switch	On
AV Search +	On	■V	
Search AV Delay	300 ms	Pace	Bipolar
Search Interval	32 cycles	Sense	Bipolar
Blanking		Safety Switch	On
A-Blank after V-Pace	125 ms	Rate Adaptive Pacing	
A-Blank after V-Sense	45 ms	Minute Ventilation	On
V-Blank after A-Pace	65 ms	Response Factor	8
Magnet Response	Pace Async	Fitness Level	Active
Noise Response	DOO	Ventilatory Threshold	120 min ⁻¹
Rate Enhancements		Ventilatory Thresh. Response	70 %
Rate Smoothing		Accelerometer	Passive
Up	Off %		
Down	Off %		
Sudden Brady Response	Off		
Setup			
Telemetry			
Enable use of ZIP™ telemetry		On	
Trending Setup			
Recording Method	30 Second Average		
Duration	25 hours		
Data Storage	Continuous		
Sleep Schedule			
Sleep Start Time	23:00 hh:mm		

Sleep Duration Leads Status Setup

●A

Pace/Sense

Pace/Sense

Daily Daily Impedance Limits Pace Threshold Intrinsic Impedance High Low Amplitude On 200 - 2000 Ω On On On On 200 - 2000 Ω Off

07 hours

2869 Software Version: 2.02 J276 Firmware Version: D_v1.02.00(1.01) © 2014

Boston Scientific Corporation
or its affiliates. All rights reserved.

Page 2 of 2

Clinician Signature: